

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011223\
 Data File : VY012169.D
 Acq On : 12 Jan 2023 11:01
 Operator : KP/MD
 Sample : VY0112SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0112SBL01

Quant Time: Jan 12 13:44:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	230300	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	390783	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	369557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	169114	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	128656	49.717	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.440%
35) Dibromofluoromethane	7.716	113	118548	46.154	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.300%
50) Toluene-d8	10.179	98	461350	45.506	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.020%
62) 4-Bromofluorobenzene	12.483	95	155784	43.961	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011223\
Data File : VY012169.D
Acq On : 12 Jan 2023 11:01
Operator : KP/MD
Sample : VY0112SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0112SBL01

Quant Time: Jan 12 13:44:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

