

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011223\
 Data File : VY012174.D
 Acq On : 12 Jan 2023 13:37
 Operator : KP/MD
 Sample : 01129-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-207

Quant Time: Jan 13 00:37:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	226954	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	387842	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	369156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	174813	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	130392	51.130	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	102.260%	
35) Dibromofluoromethane	7.716	113	117057	45.919	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	91.840%	
50) Toluene-d8	10.179	98	464277	46.142	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	92.280%	
62) 4-Bromofluorobenzene	12.483	95	155236	44.138	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	88.280%	
Target Compounds						
20) Methylene Chloride	4.716	84	35094	8.983	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011223\
Data File : VY012174.D
Acq On : 12 Jan 2023 13:37
Operator : KP/MD
Sample : 01129-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-207

Quant Time: Jan 13 00:37:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

