

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011223\
 Data File : VY012175.D
 Acq On : 12 Jan 2023 14:21
 Operator : KP/MD
 Sample : 01132-06
 Misc : 1.36g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOT-VOC-1

Quant Time: Jan 13 00:38:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	20344	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	30260	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	25548	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	9692	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	10618	46.448	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.900%
35) Dibromofluoromethane	7.716	113	8546	42.968	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.940%
50) Toluene-d8	10.179	98	34663	44.154	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.300%
62) 4-Bromofluorobenzene	12.483	95	9630	35.094	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	70.180%
Target Compounds						
16) Acetone	3.954	43	983	18.280	ug/l	Qvalue # 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011223\
Data File : VY012175.D
Acq On : 12 Jan 2023 14:21
Operator : KP/MD
Sample : 01132-06
Misc : 1.36g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOT-VOC-1

Quant Time: Jan 13 00:38:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

