


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011223\
Data File : VY012167.D
Acq On : 12 Jan 2023 09:30
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Title : SW846 8260
Last Update : Sat Jan 07 01:46:48 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.0	17601	PASS
75	95	30	60	50.2	52061	PASS
95	95	100	100	100.0	103805	PASS
96	95	5	9	6.5	6702	PASS
173	174	0.00	2	1.0	930	PASS
174	95	50	100	86.8	90109	PASS
175	174	5	9	7.6	6878	PASS
176	174	95	101	96.4	86848	PASS
177	176	5	9	6.7	5777	PASS