

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011323\
 Data File : VY012183.D
 Acq On : 13 Jan 2023 10:50
 Operator : KP/MD
 Sample : VY0113SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0113SBL01

Quant Time: Jan 16 08:16:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	199065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	331813	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	319788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	150122	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	116038	51.876	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.760%
35) Dibromofluoromethane	7.716	113	100305	45.991	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.980%
50) Toluene-d8	10.179	98	392931	45.646	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.300%
62) 4-Bromofluorobenzene	12.483	95	132575	44.060	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	88.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011323\
Data File : VY012183.D
Acq On : 13 Jan 2023 10:50
Operator : KP/MD
Sample : VY0113SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0113SBL01

Quant Time: Jan 16 08:16:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

