


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\VY011420\
 Data File : VY001189.D
 Acq On : 14 Jan 2020 08:48
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y011020S.M
 Title : SW846 8260
 Last Update : Fri Jan 10 16:01:58 2020



AutoFind: Scans 1761, 1762, 1763; Background Corrected with Scan 1753

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.0	10869	PASS
75	95	30	60	48.5	29360	PASS
95	95	100	100	100.0	60544	PASS
96	95	5	9	6.6	4020	PASS
173	174	0.00	2	0.8	367	PASS
174	95	50	100	74.4	45037	PASS
175	174	5	9	7.1	3177	PASS
176	174	95	101	98.0	44128	PASS
177	176	5	9	6.8	3014	PASS