

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\VY011520\
 Data File : VY001233.D
 Acq On : 15 Jan 2020 20:49
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 16 06:36:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y011020S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 10 16:01:58 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.384	0.426	-10.9	106	0.00
3 P	Chloromethane	0.553	0.485	12.3	87	0.00
4 C	Vinyl Chloride	0.558	0.503	9.9#	84	0.00
5 T	Bromomethane	0.300	0.339	-13.0	102	0.00
6 T	Chloroethane	0.322	0.313	2.8	91	0.00
7 T	Trichlorofluoromethane	0.776	0.833	-7.3	104	0.00
8 T	Diethyl Ether	0.302	0.358	-18.5	112	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.500	0.544	-8.8	103	0.00
10 T	Methyl Iodide	0.684	0.744	-8.8	102	0.00
11 T	Tert butyl alcohol	0.053	0.061	-15.1	121	0.01
12 CM	1,1-Dichloroethene	0.511	0.574	-12.3#	108	0.00
13 T	Acrolein	0.048	0.054	-12.5	111	0.00
14 T	Allyl chloride	0.863	0.918	-6.4	101	0.00
15 T	Acrylonitrile	0.151	0.169	-11.9	109	0.00
16 T	Acetone	0.123	0.131	-6.5	107	0.00
17 T	Carbon Disulfide	1.654	1.820	-10.0	104	0.00
18 T	Methyl Acetate	0.405	0.428	-5.7	106	0.00
19 T	Methyl tert-butyl Ether	1.414	1.615	-14.2	108	0.00
20 T	Methylene Chloride	0.683	0.729	-6.7	121	0.00
21 T	trans-1,2-Dichloroethene	0.580	0.660	-13.8	108	0.00
22 T	Diisopropyl ether	1.762	1.933	-9.7	104	0.00
23 T	Vinyl Acetate	1.127	1.222	-8.4	102	0.00
24 P	1,1-Dichloroethane	0.965	1.096	-13.6	108	0.00
25 T	2-Butanone	0.191	0.213	-11.5	108	0.00
26 T	2,2-Dichloropropane	0.887	0.883	0.5	97	0.00
27 T	cis-1,2-Dichloroethene	0.635	0.741	-16.7	110	0.00
28 T	Bromochloromethane	0.396	0.498	-25.8#	109	0.00
29 T	Tetrahydrofuran	0.131	0.143	-9.2	104	0.00
30 C	Chloroform	0.945	1.074	-13.7#	107	0.00
31 T	Cyclohexane	1.021	1.008	1.3	97	0.00
32 T	1,1,1-Trichloroethane	0.839	0.891	-6.2	100	0.00
33 S	1,2-Dichloroethane-d4	0.528	0.534	-1.1	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.293	0.275	6.1	94	0.00
36 T	1,1-Dichloropropene	0.456	0.453	0.7	100	0.00
37 T	Ethyl Acetate	0.267	0.264	1.1	102	0.00
38 T	Carbon Tetrachloride	0.418	0.401	4.1	96	0.00
39 T	Methylcyclohexane	0.618	0.589	4.7	96	0.00
40 TM	Benzene	1.330	1.408	-5.9	107	0.00
41 T	Methacrylonitrile	0.126	0.164	-30.2#	131	0.02
42 TM	1,2-Dichloroethane	0.372	0.384	-3.2	103	0.00
43 T	Isopropyl Acetate	0.511	0.520	-1.8	103	0.00
44 TM	Trichloroethene	0.347	0.346	0.3	99	0.00
45 C	1,2-Dichloropropane	0.329	0.354	-7.6#	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\VY011520\
 Data File : VY001233.D
 Acq On : 15 Jan 2020 20:49
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 16 06:36:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y011020S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 10 16:01:58 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.179	0.188	-5.0	105	0.00
47 T	Bromodichloromethane	0.437	0.452	-3.4	104	0.00
48 T	Methyl methacrylate	0.237	0.229	3.4	98	0.00
49 T	1,4-Dioxane	0.002	0.003	-50.0#	110	0.00
50 S	Toluene-d8	1.128	1.106	2.0	92	0.00
51 T	4-Methyl-2-Pentanone	0.259	0.256	1.2	100	0.00
52 CM	Toluene	0.848	0.873	-2.9#	103	0.00
53 T	t-1,3-Dichloropropene	0.489	0.492	-0.6	100	0.00
54 T	cis-1,3-Dichloropropene	0.552	0.573	-3.8	104	0.00
55 T	1,1,2-Trichloroethane	0.266	0.278	-4.5	105	0.00
56 T	Ethyl methacrylate	0.407	0.419	-2.9	103	0.00
57 T	1,3-Dichloropropane	0.457	0.485	-6.1	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.143	0.145	-1.4	112	0.00
59 T	2-Hexanone	0.182	0.179	1.6	99	0.00
60 T	Dibromochloromethane	0.306	0.311	-1.6	101	0.00
61 T	1,2-Dibromoethane	0.260	0.263	-1.2	102	0.00
62 S	4-Bromofluorobenzene	0.461	0.422	8.5	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.322	0.310	3.7	94	0.00
65 PM	Chlorobenzene	0.999	1.023	-2.4	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.344	0.349	-1.5	99	0.00
67 C	Ethyl Benzene	1.828	1.869	-2.2#	101	0.00
68 T	m/p-Xylenes	0.689	0.706	-2.5	99	0.00
69 T	o-Xylene	0.661	0.687	-3.9	102	0.00
70 T	Styrene	1.160	1.206	-4.0	101	0.00
71 P	Bromoform	0.212	0.209	1.4	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.759	3.898	-3.7	99	0.00
74 T	N-amyl acetate	1.173	1.174	-0.1	95	0.00
75 P	1,1,2,2-Tetrachloroethane	0.791	0.839	-6.1	103	0.00
76 T	1,2,3-Trichloropropane	0.535	0.561	-4.9	107	0.00
77 T	Bromobenzene	0.831	0.924	-11.2	107	0.00
78 T	n-propylbenzene	4.521	4.781	-5.8	101	0.00
79 T	2-Chlorotoluene	2.597	2.710	-4.4	100	0.00
80 T	1,3,5-Trimethylbenzene	3.136	3.269	-4.2	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.316	0.311	1.6	95	0.00
82 T	4-Chlorotoluene	2.772	2.829	-2.1	97	0.00
83 T	tert-Butylbenzene	2.635	2.680	-1.7	97	0.00
84 T	1,2,4-Trimethylbenzene	3.115	3.234	-3.8	99	0.00
85 T	sec-Butylbenzene	3.791	3.906	-3.0	98	0.00
86 T	p-Isopropyltoluene	3.292	3.356	-1.9	96	0.00
87 T	1,3-Dichlorobenzene	1.571	1.635	-4.1	100	0.00
88 T	1,4-Dichlorobenzene	1.617	1.659	-2.6	99	0.00
89 T	n-Butylbenzene	3.315	3.416	-3.0	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\VY011520\
 Data File : VY001233.D
 Acq On : 15 Jan 2020 20:49
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 16 06:36:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y011020S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 10 16:01:58 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.643	0.686	-6.7	100	0.00
91 T	1,2-Dichlorobenzene	1.460	1.514	-3.7	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.143	0.135	5.6	92	0.00
93 T	1,2,4-Trichlorobenzene	0.962	0.954	0.8	95	0.00
94 T	Hexachlorobutadiene	0.485	0.456	6.0	89	0.00
95 T	Naphthalene	2.239	2.259	-0.9	97	0.00
96 T	1,2,3-Trichlorobenzene	0.848	0.836	1.4	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6