

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011525\
 Data File : VY020831.D
 Acq On : 15 Jan 2025 12:14
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jan 16 00:58:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 01:31:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	200947	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.628	114	349203	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	296537	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	114066	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	99550	58.285	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.580%
35) Dibromofluoromethane	7.646	113	108692	49.683	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.360%
50) Toluene-d8	10.115	98	404468	53.035	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.060%
62) 4-Bromofluorobenzene	12.414	95	119901	42.574	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	85.140%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011525\
Data File : VY020831.D
Acq On : 15 Jan 2025 12:14
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jan 16 00:58:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 01:31:46 2025
Response via : Initial Calibration

