

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011525\
 Data File : VY020834.D
 Acq On : 15 Jan 2025 16:03
 Operator : SY/MD
 Sample : Q1093-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT-3407

Quant Time: Jan 16 00:59:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 01:31:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	182604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	313437	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	253527	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	88258	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	87348	56.278	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.560%
35) Dibromofluoromethane	7.646	113	98146	49.981	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.960%
50) Toluene-d8	10.115	98	356243	52.042	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	104.080%
62) 4-Bromofluorobenzene	12.414	95	98677	39.036	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	78.080%
Target Compounds						
16) Acetone	3.891	43	2383	10.797	ug/l #	Qvalue 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011525\
Data File : VY020834.D
Acq On : 15 Jan 2025 16:03
Operator : SY/MD
Sample : Q1093-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT-3407

Quant Time: Jan 16 00:59:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 01:31:46 2025
Response via : Initial Calibration

