

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011724\
 Data File : VY017007.D
 Acq On : 17 Jan 2024 12:01
 Operator : SY/MD
 Sample : VY0117SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0117SBL01

Quant Time: Jan 17 15:16:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	106432	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	225017	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.514	117	212569	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.447	152	81138	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	49428	46.988	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.980%
35) Dibromofluoromethane	7.752	113	52574	36.665	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	73.340%
50) Toluene-d8	10.209	98	197251	39.949	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	=	79.900%
62) 4-Bromofluorobenzene	12.508	95	62833	37.717	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	=	75.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011724\
Data File : VY017007.D
Acq On : 17 Jan 2024 12:01
Operator : SY/MD
Sample : VY0117SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0117SBL01

Quant Time: Jan 17 15:16:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

