

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011724\
 Data File : VY017009.D
 Acq On : 17 Jan 2024 14:38
 Operator : SY/MD
 Sample : P1117-05
 Misc : 3.60g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S1-TOTAL-VOC-GRAB-1

Quant Time: Jan 17 15:16:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	36782	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.710	114	63679	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.514	117	47522	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.453	152	13664	50.000	ug/l	0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	25205	69.332	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	138.660%
35) Dibromofluoromethane	7.746	113	24365	60.044	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	120.080%
50) Toluene-d8	10.203	98	68705	49.169	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.340%
62) 4-Bromofluorobenzene	12.508	95	17617	37.368	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	=	74.740%
Target Compounds						
20) Methylene Chloride	4.753	84	5446	5.123	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011724\
Data File : VY017009.D
Acq On : 17 Jan 2024 14:38
Operator : SY/MD
Sample : P1117-05
Misc : 3.60g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S1-TOTAL-VOC-GRAB-1

Quant Time: Jan 17 15:16:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

