

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011724\
 Data File : VY017011.D
 Acq On : 17 Jan 2024 15:23
 Operator : SY/MD
 Sample : P1117-12
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S2-TOTAL-VOC-GRAB-8

Quant Time: Jan 17 15:29:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	3126	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.716	114	3738	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.514	117	1893	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.447	152	427	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	1176	38.063	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	76.120%
35) Dibromofluoromethane	7.740	113	1055	44.291	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.580%
50) Toluene-d8	10.203	98	2965	36.148	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	72.300%
62) 4-Bromofluorobenzene	12.508	95	665	24.030	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	=	48.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011724\
Data File : VY017011.D
Acq On : 17 Jan 2024 15:23
Operator : SY/MD
Sample : P1117-12
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S2-TOTAL-VOC-GRAB-8

Quant Time: Jan 17 15:29:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

