

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011724\
 Data File : VY017018.D
 Acq On : 17 Jan 2024 18:05
 Operator : SY/MD
 Sample : P1140-03RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MR-309-37-39RE

Quant Time: Jan 18 00:46:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	4522	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	7988	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.520	117	6563	50.000	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	13.446	152	1821	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	5408	121.001	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	= 242.000%#	
35) Dibromofluoromethane	7.746	113	5357	105.241	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	= 210.480%#	
50) Toluene-d8	10.209	98	14909	85.056	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	= 170.120%#	
62) 4-Bromofluorobenzene	12.507	95	3865	65.355	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	= 130.720%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011724\
Data File : VY017018.D
Acq On : 17 Jan 2024 18:05
Operator : SY/MD
Sample : P1140-03RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MR-309-37-39RE

Quant Time: Jan 18 00:46:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

