

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011724\
 Data File : VY017022.D
 Acq On : 17 Jan 2024 19:38
 Operator : SY/MD
 Sample : P1155-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BB-TESTPITS

Quant Time: Jan 18 00:47:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.825	168	100190	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	219168	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.520	117	209447	50.000	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	13.453	152	82146	50.000	ug/l	0.02

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.179	65	48941	49.423	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.840%
35) Dibromofluoromethane	7.758	113	53108	38.026	ug/l	0.03
Spiked Amount	50.000	Range	54 - 147	Recovery	=	76.060%
50) Toluene-d8	10.209	98	190822	39.678	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	=	79.360%
62) 4-Bromofluorobenzene	12.508	95	61157	37.691	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	=	75.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011724\
Data File : VY017022.D
Acq On : 17 Jan 2024 19:38
Operator : SY/MD
Sample : P1155-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BB-TESTPITS

Quant Time: Jan 18 00:47:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

