



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011724\
Data File : VY017005.D
Acq On : 17 Jan 2024 08:47
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Title : SW846 8260
Last Update : Tue Jan 02 14:25:32 2024



AutoFind: Scans 1753, 1754, 1755; Background Corrected with Scan 1745

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	14609	PASS
75	95	30	60	52.9	37264	PASS
95	95	100	100	100.0	70437	PASS
96	95	5	9	6.8	4818	PASS
173	174	0.00	2	1.3	703	PASS
174	95	50	100	75.9	53461	PASS
175	174	5	9	7.6	4075	PASS
176	174	95	101	96.3	51483	PASS
177	176	5	9	6.4	3284	PASS