

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011725\
 Data File : VY020848.D
 Acq On : 17 Jan 2025 12:26
 Operator : SY/MD
 Sample : Q1118-01
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-01-1-16-2025

Quant Time: Jan 17 23:24:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 01:31:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	195637	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	342704	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	293504	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	116598	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	97083	58.384	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.760%
35) Dibromofluoromethane	7.640	113	108500	50.535	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.080%
50) Toluene-d8	10.109	98	397698	53.136	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.280%
62) 4-Bromofluorobenzene	12.408	95	122460	44.307	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	88.620%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011725\
Data File : VY020848.D
Acq On : 17 Jan 2025 12:26
Operator : SY/MD
Sample : Q1118-01
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-01-1-16-2025

Quant Time: Jan 17 23:24:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 01:31:46 2025
Response via : Initial Calibration

