

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011725\
 Data File : VY020852.D
 Acq On : 17 Jan 2025 14:00
 Operator : SY/MD
 Sample : Q1115-06
 Misc : 3.38g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-214-VOC

Quant Time: Jan 17 23:25:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 01:31:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	178344	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	305063	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	269564	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	108194	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	90323	59.585	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 119.180%			
35) Dibromofluoromethane	7.640	113	98943	51.770	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 103.540%			
50) Toluene-d8	10.109	98	357761	53.698	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 107.400%			
62) 4-Bromofluorobenzene	12.408	95	111454	45.300	ug/l	0.00
Spiked Amount 50.000	Range 29 - 146		Recovery = 90.600%			
Target Compounds						
51) 4-Methyl-2-Pentanone	9.999	43	1546	1.568	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011725\
Data File : VY020852.D
Acq On : 17 Jan 2025 14:00
Operator : SY/MD
Sample : Q1115-06
Misc : 3.38g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-214-VOC

Quant Time: Jan 17 23:25:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 01:31:46 2025
Response via : Initial Calibration

