



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011725\
Data File : VY020841.D
Acq On : 17 Jan 2025 08:47
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Title : SW846 8260
Last Update : Tue Jan 07 01:31:46 2025



AutoFind: Scans 1754, 1755, 1756; Background Corrected with Scan 1746

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	16405	PASS
75	95	30	60	49.7	50795	PASS
95	95	100	100	100.0	102176	PASS
96	95	5	9	6.8	6982	PASS
173	174	0.00	2	0.8	675	PASS
174	95	50	100	86.7	88557	PASS
175	174	5	9	7.7	6844	PASS
176	174	95	101	96.4	85344	PASS
177	176	5	9	7.1	6067	PASS