

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011824\
 Data File : VY017030.D
 Acq On : 18 Jan 2024 11:03
 Operator : SY/MD
 Sample : VY0118SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0118SBL01

Quant Time: Jan 19 02:24:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.826	168	109576	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	234729	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.514	117	222419	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.453	152	83824	50.000	ug/l	0.02

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	53683	49.568	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.140%
35) Dibromofluoromethane	7.758	113	56014	37.448	ug/l	0.03
Spiked Amount	50.000	Range	54 - 147	Recovery	=	74.900%
50) Toluene-d8	10.209	98	204488	39.701	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	=	79.400%
62) 4-Bromofluorobenzene	12.508	95	65259	37.553	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	=	75.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011824\
Data File : VY017030.D
Acq On : 18 Jan 2024 11:03
Operator : SY/MD
Sample : VY0118SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0118SBL01

Quant Time: Jan 19 02:24:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

