

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
 Data File : VY012260.D
 Acq On : 19 Jan 2023 13:03
 Operator : KP/MD
 Sample : 01198-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PL-02-011823

Quant Time: Jan 20 03:23:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	189287	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	310079	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	298684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	137504	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	107515	70.107	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	140.220%
35) Dibromofluoromethane	7.716	113	92405	56.280	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.560%
50) Toluene-d8	10.179	98	369703	66.410	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	132.820%
62) 4-Bromofluorobenzene	12.483	95	122804	53.058	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
Data File : VY012260.D
Acq On : 19 Jan 2023 13:03
Operator : KP/MD
Sample : 01198-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PL-02-011823

Quant Time: Jan 20 03:23:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

