

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
 Data File : VY012264.D
 Acq On : 19 Jan 2023 14:37
 Operator : KP/MD
 Sample : 01214-10
 Misc : 5.75g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TJRC-011823-B3-20-30

Quant Time: Jan 20 03:24:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	180477	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	298743	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	288128	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	144227	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	107817	73.957	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 147.920%	
35) Dibromofluoromethane	7.715	113	89228	56.408	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 112.820%	
50) Toluene-d8	10.178	98	353813	65.936	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 131.880%	
62) 4-Bromofluorobenzene	12.483	95	125764	56.399	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 112.800%	
Target Compounds						
30) Chloroform	7.508	83	5101	1.525	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
Data File : VY012264.D
Acq On : 19 Jan 2023 14:37
Operator : KP/MD
Sample : 01214-10
Misc : 5.75g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TJRC-011823-B3-20-30

Quant Time: Jan 20 03:24:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

