

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
 Data File : VY012265.D
 Acq On : 19 Jan 2023 15:00
 Operator : KP/MD
 Sample : 01214-13
 Misc : 7.30g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TJRC-011823-B3-30-40

Quant Time: Jan 20 03:24:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	203621	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	332854	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	331466	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	160474	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	106914	64.485	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.960%	
35) Dibromofluoromethane	7.716	113	95467	54.167	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.340%	
50) Toluene-d8	10.179	98	401338	67.213	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 134.420%	
62) 4-Bromofluorobenzene	12.483	95	141626	57.003	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 114.000%	
Target Compounds						
					Qvalue	
30) Chloroform	7.508	83	13198	3.497	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
Data File : VY012265.D
Acq On : 19 Jan 2023 15:00
Operator : KP/MD
Sample : 01214-13
Misc : 7.30g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TJRC-011823-B3-30-40

Quant Time: Jan 20 03:24:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

