

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
 Data File : VY012271.D
 Acq On : 19 Jan 2023 17:21
 Operator : KP/MD
 Sample : 01190-11
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PE-10

Quant Time: Jan 20 03:25:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	1241	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	2238	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	1515	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	243	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	914	92.171	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	184.340%#
35) Dibromofluoromethane	7.716	113	288	24.303	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	48.600%#
50) Toluene-d8	10.185	98	2350	57.921	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.840%
62) 4-Bromofluorobenzene	12.483	95	331	19.814	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	39.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
Data File : VY012271.D
Acq On : 19 Jan 2023 17:21
Operator : KP/MD
Sample : 01190-11
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PE-10

Quant Time: Jan 20 03:25:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

