

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
 Data File : VY012273.D
 Acq On : 19 Jan 2023 18:08
 Operator : KP/MD
 Sample : 01190-05
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PE-4

Quant Time: Jan 20 03:25:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	95987	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	165002	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	191231	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	97634	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	63317	82.106	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	164.220%#
35) Dibromofluoromethane	7.722	113	51106	58.495	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.980%
50) Toluene-d8	10.179	98	213422	72.447	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	144.900%#
62) 4-Bromofluorobenzene	12.483	95	87878	71.351	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	142.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
Data File : VY012273.D
Acq On : 19 Jan 2023 18:08
Operator : KP/MD
Sample : 01190-05
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PE-4

Quant Time: Jan 20 03:25:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

