

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
 Data File : VY012276.D
 Acq On : 19 Jan 2023 19:18
 Operator : KP/MD
 Sample : 01217-01
 Misc : 5.39g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BORING-4

Quant Time: Jan 20 03:26:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	271872	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	458035	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	441233	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	211274	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	149480	67.726	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.460%
35) Dibromofluoromethane	7.716	113	138680	57.181	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.360%
50) Toluene-d8	10.179	98	535388	65.013	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	130.020%
62) 4-Bromofluorobenzene	12.483	95	192151	56.202	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
Data File : VY012276.D
Acq On : 19 Jan 2023 19:18
Operator : KP/MD
Sample : 01217-01
Misc : 5.39g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BORING-4

Quant Time: Jan 20 03:26:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

