

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
 Data File : VY012253.D
 Acq On : 19 Jan 2023 09:48
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 20 03:21:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	119	0.00
2 T	Dichlorodifluoromethane	0.437	0.378	13.5	121	0.00
3 P	Chloromethane	0.403	0.379	6.0	122	0.00
4 C	Vinyl Chloride	0.471	0.449	4.7#	123	0.00
5 T	Bromomethane	0.364	0.346	4.9	123	0.00
6 T	Chloroethane	0.307	0.296	3.6	122	0.00
7 T	Trichlorofluoromethane	0.919	0.854	7.1	120	0.00
8 T	Diethyl Ether	0.273	0.267	2.2	122	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.447	0.443	0.9	129	0.00
10 T	Methyl Iodide	0.680	0.621	8.7	108	0.00
11 T	Tert butyl alcohol	0.057	0.047	17.5	122	0.00
12 CM	1,1-Dichloroethene	0.459	0.453	1.3#	123	0.00
13 T	Acrolein	0.023	0.014	39.1#	106	0.00
14 T	Allyl chloride	0.568	0.589	-3.7	129	0.00
15 T	Acrylonitrile	0.119	0.120	-0.8	122	0.00
16 T	Acetone	0.139	0.153	-10.1	118	0.00
17 T	Carbon Disulfide	1.420	1.400	1.4	124	0.00
18 T	Methyl Acetate	0.312	0.314	-0.6	124	0.00
19 T	Methyl tert-butyl Ether	1.280	1.258	1.7	117	0.00
20 T	Methylene Chloride	0.762	0.520	31.8#	108	0.00
21 T	trans-1,2-Dichloroethene	0.512	0.517	-1.0	124	0.00
22 T	Diisopropyl ether	1.132	1.191	-5.2	129	0.00
23 T	Vinyl Acetate	0.705	0.722	-2.4	122	0.00
24 P	1,1-Dichloroethane	0.785	0.807	-2.8	126	0.00
25 T	2-Butanone	0.154	0.164	-6.5	124	0.00
26 T	2,2-Dichloropropane	0.852	0.829	2.7	121	0.00
27 T	cis-1,2-Dichloroethene	0.563	0.564	-0.2	122	0.00
28 T	Bromochloromethane	0.171	0.163	4.7	121	0.00
29 T	Tetrahydrofuran	0.084	0.086	-2.4	120	0.00
30 C	Chloroform	0.927	0.916	1.2#	121	0.00
31 T	Cyclohexane	0.729	0.704	3.4	132	0.00
32 T	1,1,1-Trichloroethane	0.918	0.878	4.4	118	0.00
33 S	1,2-Dichloroethane-d4	0.465	0.397	14.6	115	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	117	0.00
35 S	Dibromofluoromethane	0.265	0.255	3.8	124	0.00
36 T	1,1-Dichloropropene	0.450	0.451	-0.2	124	0.00
37 T	Ethyl Acetate	0.198	0.199	-0.5	119	0.00
38 T	Carbon Tetrachloride	0.581	0.555	4.5	118	0.00
39 T	Methylcyclohexane	0.550	0.559	-1.6	127	0.00
40 TM	Benzene	1.299	1.312	-1.0	124	0.00
41 T	Methacrylonitrile	0.117	0.124	-6.0	121	-0.02
42 TM	1,2-Dichloroethane	0.413	0.400	3.1	115	0.00
43 T	Isopropyl Acetate	0.369	0.370	-0.3	118	0.00
44 TM	Trichloroethene	0.388	0.389	-0.3	123	0.00
45 C	1,2-Dichloropropane	0.271	0.290	-7.0#	130	0.00
46 T	Dibromomethane	0.181	0.183	-1.1	120	0.00
47 T	Bromodichloromethane	0.457	0.457	0.0	120	0.00
48 T	Methyl methacrylate	0.170	0.170	0.0	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
 Data File : VY012253.D
 Acq On : 19 Jan 2023 09:48
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 20 03:21:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	113	0.00
50 S	Toluene-d8	1.044	0.920	11.9	121	0.00
51 T	4-Methyl-2-Pentanone	0.197	0.200	-1.5	118	0.00
52 CM	Toluene	0.865	0.872	-0.8#	122	0.00
53 T	t-1,3-Dichloropropene	0.478	0.477	0.2	118	0.00
54 T	cis-1,3-Dichloropropene	0.514	0.522	-1.6	121	0.00
55 T	1,1,2-Trichloroethane	0.258	0.262	-1.6	121	0.00
56 T	Ethyl methacrylate	0.338	0.352	-4.1	119	0.00
57 T	1,3-Dichloropropane	0.421	0.433	-2.9	122	0.00
58 T	2-Chloroethyl Vinyl ether	0.112	0.106	5.4	115	0.00
59 T	2-Hexanone	0.150	0.159	-6.0	118	0.00
60 T	Dibromochloromethane	0.342	0.341	0.3	116	0.00
61 T	1,2-Dibromoethane	0.249	0.250	-0.4	119	0.00
62 S	4-Bromofluorobenzene	0.373	0.347	7.0	122	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	116	0.00
64 T	Tetrachloroethene	0.456	0.488	-7.0	126	0.00
65 PM	Chlorobenzene	1.023	1.020	0.3	120	0.00
66 T	1,1,1,2-Tetrachloroethane	0.386	0.395	-2.3	119	0.00
67 C	Ethyl Benzene	1.828	1.850	-1.2#	120	0.00
68 T	m/p-Xylenes	0.723	0.734	-1.5	120	0.00
69 T	o-Xylene	0.690	0.695	-0.7	119	0.00
70 T	Styrene	1.154	1.182	-2.4	119	0.00
71 P	Bromoform	0.247	0.246	0.4	115	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
73 T	Isopropylbenzene	3.605	3.751	-4.0	120	0.00
74 T	N-amyl acetate	0.707	0.718	-1.6	112	0.00
75 P	1,1,2,2-Tetrachloroethane	0.568	0.577	-1.6	115	0.00
76 T	1,2,3-Trichloropropane	0.460	0.431	6.3	95	0.00
77 T	Bromobenzene	0.868	0.892	-2.8	117	0.00
78 T	n-propylbenzene	4.219	4.338	-2.8	120	0.00
79 T	2-Chlorotoluene	2.385	2.445	-2.5	119	0.00
80 T	1,3,5-Trimethylbenzene	3.134	3.215	-2.6	119	0.00
81 T	trans-1,4-Dichloro-2-butene	0.223	0.233	-4.5	115	0.00
82 T	4-Chlorotoluene	2.506	2.523	-0.7	118	0.00
83 T	tert-Butylbenzene	2.732	2.786	-2.0	121	0.00
84 T	1,2,4-Trimethylbenzene	3.091	3.168	-2.5	117	0.00
85 T	sec-Butylbenzene	3.889	3.945	-1.4	119	0.00
86 T	p-Isopropyltoluene	3.398	3.428	-0.9	118	0.00
87 T	1,3-Dichlorobenzene	1.725	1.728	-0.2	116	0.00
88 T	1,4-Dichlorobenzene	1.721	1.712	0.5	116	0.00
89 T	n-Butylbenzene	2.945	3.008	-2.1	120	0.00
90 T	Hexachloroethane	0.613	0.615	-0.3	118	0.00
91 T	1,2-Dichlorobenzene	1.540	1.515	1.6	114	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.117	0.113	3.4	109	0.00
93 T	1,2,4-Trichlorobenzene	0.987	0.948	4.0	114	0.00
94 T	Hexachlorobutadiene	0.594	0.579	2.5	118	0.00
95 T	Naphthalene	1.947	1.842	5.4	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011923\
Data File : VY012253.D
Acq On : 19 Jan 2023 09:48
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jan 20 03:21:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.849	0.797	6.1	112	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6