

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011924\
 Data File : VY017058.D
 Acq On : 19 Jan 2024 16:09
 Operator : SY/MD
 Sample : P1180-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-01-011824

Quant Time: Jan 19 22:23:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.819	168	96654	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.716	114	206086	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.508	117	201470	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.446	152	83215	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	42475	44.463	ug/l	0.01
Spiked Amount 50.000	Range 50 - 163		Recovery =	88.920%		
35) Dibromofluoromethane	7.746	113	48401	36.856	ug/l	0.02
Spiked Amount 50.000	Range 54 - 147		Recovery =	73.720%		
50) Toluene-d8	10.203	98	182962	40.458	ug/l	0.01
Spiked Amount 50.000	Range 58 - 134		Recovery =	80.920%		
62) 4-Bromofluorobenzene	12.501	95	62235	40.790	ug/l	0.01
Spiked Amount 50.000	Range 30 - 143		Recovery =	81.580%		
Target Compounds						
16) Acetone	3.979	43	2921	Below Cal	#	87
20) Methylene Chloride	4.759	84	11394	2.510	ug/l	97
25) 2-Butanone	7.027	43	313	Below Cal	#	48
31) Cyclohexane	7.825	56	6822	4.161	ug/l	# 75
36) 1,1-Dichloropropene	7.819	75	8113	4.239	ug/l	# 49
38) Carbon Tetrachloride	7.819	117	9579	4.926	ug/l	# 16
39) Methylcyclohexane	9.203	83	6291	2.924	ug/l	# 85
40) Benzene	8.191	78	23622	4.114	ug/l	98
51) 4-Methyl-2-Pentanone	10.203	43	968	1.093	ug/l	# 1
52) Toluene	10.264	92	6343	1.831	ug/l	99
55) 1,1,2-Trichloroethane	10.642	97	1237	1.122	ug/l	# 15
67) Ethyl Benzene	11.611	91	21738	2.877	ug/l	97
68) m/p-Xylenes	11.721	106	18378	6.626	ug/l	96
69) o-Xylene	12.050	106	29338	11.283	ug/l	99
73) Isopropylbenzene	12.349	105	8531	1.255	ug/l	98
74) N-amyl acetate	12.142	43	1511	1.107	ug/l	# 71
76) 1,2,3-Trichloropropane	12.501	75	33806	37.436	ug/l	# 100
78) n-propylbenzene	12.690	91	25554	3.093	ug/l	99
79) 2-Chlorotoluene	12.751	91	11250	2.456	ug/l	# 41
80) 1,3,5-Trimethylbenzene	12.831	105	60898	11.331	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.501	75	33806	85.008	ug/l	# 12
82) 4-Chlorotoluene	12.831	91	6677	1.427	ug/l	# 41
83) tert-Butylbenzene	13.142	119	12254	2.589	ug/l	# 60
84) 1,2,4-Trimethylbenzene	13.142	105	100384	18.967	ug/l	98
85) sec-Butylbenzene	13.270	105	14350	2.046	ug/l	87
86) p-Isopropyltoluene	13.385	119	12255	2.117	ug/l	98
89) n-Butylbenzene	13.715	91	16002	2.875	ug/l	# 72
90) Hexachloroethane	13.989	117	5840	4.695	ug/l	# 17
92) 1,2-Dibromo-3-Chloropr...	14.349	75	465	2.678	ug/l	# 1
95) Naphthalene	15.257	128	20374	10.178	ug/l	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011924\
Data File : VY017058.D
Acq On : 19 Jan 2024 16:09
Operator : SY/MD
Sample : P1180-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-01-011824

Quant Time: Jan 19 22:23:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.819 min Scan# 91
Delta R.T. 0.018 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09




Instrument :
MSVOA_Y
ClientSampleId :
EO-01-011824


Tgt Ion:168 Resp: 96654
Ion Ratio Lower Upper
168 100
99 55.3 44.2 66.4



#16
Acetone
Concen: Below Cal
RT: 3.979 min Scan# 354
Delta R.T. 0.019 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09

Tgt Ion: 43 Resp: 2921
Ion Ratio Lower Upper
43 100
58 25.7 26.6 39.8#





#20

Methylene Chloride

Concen: 2.510 ug/l

RT: 4.759 min Scan# 41

Delta R.T. 0.025 min

Lab File: VY017058.D

Acq: 19 Jan 2024 16:09

Instrument :

MSVOA_Y

ClientSampleId :

EO-01-011824

Tgt Ion: 84 Resp: 11394





Ion Ratio Lower Upper

84 100

49 122.8 95.2 142.8

51 35.0 29.9 44.9

86 63.5 52.2 78.4



#25

2-Butanone

Concen: Below Cal

RT: 7.027 min Scan# 854

Delta R.T. 0.031 min

Lab File: VY017058.D




Acq: 19 Jan 2024 16:09


Tgt Ion: 43 Resp: 313

Ion Ratio Lower Upper

43 100

72 0.0 21.4 32.2#









#31
Cyclohexane
Concen: 4.161 ug/l
RT: 7.825 min Scan# 91
Delta R.T. 0.024 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09




Instrument :
MSVOA_Y
ClientSampleId :
EO-01-011824


Tgt Ion: 56 Resp: 6822
Ion Ratio Lower Upper
56 100
69 58.9 27.0 40.6#
84 106.8 71.8 107.6



#33
1,2-Dichloroethane-d4
Concen: 44.463 ug/l
RT: 8.167 min Scan# 1041
Delta R.T. 0.012 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09

Tgt Ion: 65 Resp: 42475
Ion Ratio Lower Upper
65 100
67 54.5 0.0 109.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.716 min Scan# 1128

Delta R.T. 0.013 min

Lab File: VY017058.D

Acq: 19 Jan 2024 16:09

Instrument :

MSVOA_Y

ClientSampleId :

EO-01-011824





Tgt Ion:114 Resp: 206086

Ion Ratio Lower Upper

114 100

63 19.6 0.0 37.4

88 15.1 0.0 29.0



#35

Dibromofluoromethane

Concen: 36.856 ug/l

RT: 7.746 min Scan# 972

Delta R.T. 0.018 min

Lab File: VY017058.D

Acq: 19 Jan 2024 16:09




Tgt Ion:113 Resp: 48401


Ion Ratio Lower Upper

113 100

111 102.5 83.8 125.6

192 18.0 14.5 21.7









#36
1,1-Dichloropropene
Concen: 4.239 ug/l
RT: 7.819 min Scan# 984
Delta R.T. -0.110 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09




Instrument :
MSVOA_Y
ClientSampleId :
EO-01-011824


Tgt Ion: 75 Resp: 8113
Ion Ratio Lower Upper
75 100
110 8.5 17.4 52.3#
77 0.0 24.7 37.1#



#38
Carbon Tetrachloride
Concen: 4.926 ug/l
RT: 7.819 min Scan# 984
Delta R.T. -0.098 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09

Tgt Ion: 117 Resp: 9579
Ion Ratio Lower Upper
117 100
119 5.0 76.0 114.0#
121 0.0 24.2 36.2#





#39

Methylcyclohexane

Concen: 2.924 ug/l

RT: 9.203 min Scan# 11

Delta R.T. 0.012 min

Lab File: VY017058.D

Acq: 19 Jan 2024 16:09

Instrument :

MSVOA_Y

ClientSampleId :

EO-01-011824





Tgt Ion: 83 Resp: 6291

Ion Ratio Lower Upper

83 100

55 62.2 59.7 89.5

98 57.5 38.0 57.0#



#40

Benzene

Concen: 4.114 ug/l

RT: 8.191 min Scan# 1045

Delta R.T. 0.018 min

Lab File: VY017058.D




Acq: 19 Jan 2024 16:09


Tgt Ion: 78 Resp: 23622

Ion Ratio Lower Upper

78 100

77 24.2 18.8 28.2





#50

Toluene-d8

Concen: 40.458 ug/l

RT: 10.203 min Scan# 1375

Delta R.T. 0.012 min

Lab File: VY017058.D

Acq: 19 Jan 2024 16:09

Instrument :

MSVOA_Y

ClientSampleId :





EO-01-011824

Tgt Ion: 98 Resp: 182962

Ion Ratio Lower Upper

98 100

100 64.9 52.3 78.5



#51

4-Methyl-2-Pentanone

Concen: 1.093 ug/l

RT: 10.203 min Scan# 1375

Delta R.T. 0.122 min

Lab File: VY017058.D




Acq: 19 Jan 2024 16:09


Tgt Ion: 43 Resp: 968

Ion Ratio Lower Upper

43 100

58 168.4 32.2 48.2#





#52

Toluene

Concen: 1.831 ug/l

RT: 10.264 min Scan# 11

Delta R.T. 0.012 min

Lab File: VY017058.D

Acq: 19 Jan 2024 16:09

Instrument :

MSVOA_Y

ClientSampleId :





EO-01-011824

Tgt Ion: 92 Resp: 6343

Ion Ratio Lower Upper

92 100

91 170.3 137.6 206.4



#55

1,1,2-Trichloroethane

Concen: 1.122 ug/l

RT: 10.642 min Scan# 1447

Delta R.T. -0.012 min

Lab File: VY017058.D

Acq: 19 Jan 2024 16:09

Tgt Ion: 97 Resp: 1237




Ion Ratio Lower Upper


97 100

83 0.0 69.4 104.2#

85 0.0 43.9 65.9#

99 0.0 48.6 73.0#









#62
4-Bromofluorobenzene
Concen: 40.790 ug/l
RT: 12.501 min Scan# 1752
Delta R.T. 0.012 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09




Instrument :
MSVOA_Y
ClientSampleId :
EO-01-011824


Tgt Ion:	95	Resp:	62235
Ion	Ratio	Lower	Upper
95	100		
174	79.4	0.0	160.0
176	76.3	0.0	151.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.508 min Scan# 1589
Delta R.T. 0.006 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09

Tgt Ion:	117	Resp:	201470
Ion	Ratio	Lower	Upper
117	100		
82	55.3	43.8	65.8
119	32.4	26.5	39.7









#67
Ethyl Benzene
Concen: 2.877 ug/l
RT: 11.611 min Scan# 1606
Delta R.T. 0.012 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09




Instrument :
MSVOA_Y
ClientSampleId :
EO-01-011824


Tgt Ion: 91 Resp: 21738
Ion Ratio Lower Upper
91 100
106 29.7 25.0 37.4



#68
m/p-Xylenes
Concen: 6.626 ug/l
RT: 11.721 min Scan# 1624
Delta R.T. 0.012 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09

Tgt Ion: 106 Resp: 18378
Ion Ratio Lower Upper
106 100
91 207.7 160.7 241.1









#69
o-Xylene
Concen: 11.283 ug/l
RT: 12.050 min Scan# 1
Delta R.T. 0.012 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09




Instrument :
MSVOA_Y
ClientSampleId :
EO-01-011824


Tgt Ion:106 Resp: 29338
Ion Ratio Lower Upper
106 100
91 211.8 106.5 319.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. 0.012 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09

Tgt Ion:152 Resp: 83215
Ion Ratio Lower Upper
152 100
115 64.5 30.0 90.0
150 156.4 0.0 344.6









#73
Isopropylbenzene
Concen: 1.255 ug/l
RT: 12.349 min Scan# 1727
Delta R.T. 0.012 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09




Instrument :
MSVOA_Y
ClientSampleId :
EO-01-011824


Tgt Ion:105 Resp: 8531
Ion Ratio Lower Upper
105 100
120 27.0 13.0 39.0



#74
N-ethyl acetate
Concen: 1.107 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. -0.012 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09

Tgt Ion: 43 Resp: 1511
Ion Ratio Lower Upper
43 100
70 29.9 36.6 55.0#
55 18.7 22.2 33.4#
61 0.0 20.2 30.4#





#76

1,2,3-Trichloropropane

Concen: 37.436 ug/l

RT: 12.501 min Scan# 1752

Delta R.T. -0.141 min

Lab File: VY017058.D

Acq: 19 Jan 2024 16:09

Instrument :

MSVOA_Y

ClientSampleId :





EO-01-011824

Tgt Ion: 75 Resp: 33806

Ion Ratio Lower Upper

75 100

77 1.7 0.0 0.0#



#78

n-propylbenzene

Concen: 3.093 ug/l

RT: 12.690 min Scan# 1783

Delta R.T. 0.012 min

Lab File: VY017058.D




Acq: 19 Jan 2024 16:09


Tgt Ion: 91 Resp: 25554

Ion Ratio Lower Upper

91 100

120 22.3 11.0 33.0









#79
2-Chlorotoluene
Concen: 2.456 ug/l
RT: 12.751 min Scan# 1793
Delta R.T. -0.012 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09




Instrument :
MSVOA_Y
ClientSampleId :
EO-01-011824


Tgt Ion: 91 Resp: 11250
Ion Ratio Lower Upper
91 100
126 0.0 17.0 51.0#



#80
1,3,5-Trimethylbenzene
Concen: 11.331 ug/l
RT: 12.831 min Scan# 1806
Delta R.T. 0.013 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09

Tgt Ion:105 Resp: 60898
Ion Ratio Lower Upper
105 100
120 48.0 24.3 72.8









#81
trans-1,4-Dichloro-2-butene
Concen: 85.008 ug/l
RT: 12.501 min Scan# 11
Delta R.T. 0.116 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09




Instrument :
MSVOA_Y
ClientSampleId :
EO-01-011824


Tgt Ion: 75 Resp: 33806
Ion Ratio Lower Upper
75 100
53 0.2 74.0 111.0#
89 0.0 36.5 54.7#



#82
4-Chlorotoluene
Concen: 1.427 ug/l
RT: 12.831 min Scan# 1806
Delta R.T. -0.030 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09

Tgt Ion: 91 Resp: 6677
Ion Ratio Lower Upper
91 100
126 0.0 17.1 51.2#





#83

tert-Butylbenzene

Concen: 2.589 ug/l

RT: 13.142 min Scan# 11

Delta R.T. 0.055 min

Lab File: VY017058.D

Acq: 19 Jan 2024 16:09

Instrument :

MSVOA_Y

ClientSampleId :

EO-01-011824





Tgt Ion:119 Resp: 12254

Ion Ratio Lower Upper

119 100

91 90.1 31.5 94.5

134 0.0 13.1 39.3#



#84

1,2,4-Trimethylbenzene

Concen: 18.967 ug/l

RT: 13.142 min Scan# 1857

Delta R.T. 0.013 min

Lab File: VY017058.D




Acq: 19 Jan 2024 16:09


Tgt Ion:105 Resp: 100384

Ion Ratio Lower Upper

105 100

120 45.8 22.1 66.3









#85
 sec-Butylbenzene
 Concen: 2.046 ug/l
 RT: 13.270 min Scan# 1878
 Delta R.T. 0.007 min
 Lab File: VY017058.D
 Acq: 19 Jan 2024 16:09




Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-01-011824


Tgt Ion:105 Resp: 14350
 Ion Ratio Lower Upper
 105 100
 134 25.3 9.7 28.9



#86
 p-Isopropyltoluene
 Concen: 2.117 ug/l
 RT: 13.385 min Scan# 1897
 Delta R.T. 0.006 min
 Lab File: VY017058.D
 Acq: 19 Jan 2024 16:09

Tgt Ion:119 Resp: 12255
 Ion Ratio Lower Upper
 119 100
 134 27.0 13.1 39.1
 91 24.6 11.7 35.0





#89

n-Butylbenzene

Concen: 2.875 ug/l

RT: 13.715 min Scan# 1951

Delta R.T. 0.013 min

Lab File: VY017058.D

Acq: 19 Jan 2024 16:09





Instrument :

MSVOA_Y

ClientSampleId :

EO-01-011824

Tgt Ion: 91	Resp: 16002
Ion Ratio	Lower Upper
91	100
92	65.4 26.4 79.2
134	0.0 12.3 36.8#



#90

Hexachloroethane

Concen: 4.695 ug/l




RT: 13.989 min Scan# 1996


Delta R.T. 0.018 min

Lab File: VY017058.D

Acq: 19 Jan 2024 16:09

Tgt Ion: 117	Resp: 5840
Ion Ratio	Lower Upper
117	100
201	0.0 33.4 100.1#









#92
1,2-Dibromo-3-Chloropropane
Concen: 2.678 ug/l
RT: 14.349 min Scan# 2055
Delta R.T. -0.018 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09

Instrument :
MSVOA_Y
ClientSampleId :
EO-01-011824

Tgt Ion: 75 Resp: 465
Ion Ratio Lower Upper
75 100
155 0.0 42.1 126.3#
157 0.0 55.8 167.4#



#95
Naphthalene
Concen: 10.178 ug/l
RT: 15.257 min Scan# 2204
Delta R.T. 0.012 min
Lab File: VY017058.D
Acq: 19 Jan 2024 16:09

Tgt Ion:128 Resp: 20374
Ion Ratio Lower Upper
128 100
127 14.9 10.2 15.4
129 9.9 8.6 12.8

