

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011924\
 Data File : VY017053.D
 Acq On : 19 Jan 2024 14:12
 Operator : SY/MD
 Sample : P1184-09
 Misc : 3.68g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-5-19.5-20.0

Quant Time: Jan 19 22:22:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	99023	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.716	114	215017	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.514	117	205813	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	76944	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	46708	47.724	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.440%
35) Dibromofluoromethane	7.746	113	51635	37.685	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	75.380%
50) Toluene-d8	10.203	98	186989	39.631	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	79.260%
62) 4-Bromofluorobenzene	12.501	95	58159	36.535	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	73.080%
Target Compounds						
20) Methylene Chloride	4.747	84	17317	7.458	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011924\
Data File : VY017053.D
Acq On : 19 Jan 2024 14:12
Operator : SY/MD
Sample : P1184-09
Misc : 3.68g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-5-19.5-20.0

Quant Time: Jan 19 22:22:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

