

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011924\
 Data File : VY017054.D
 Acq On : 19 Jan 2024 14:35
 Operator : SY/MD
 Sample : P1185-01
 Misc : 3.44g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-7-19.5-20.0

Quant Time: Jan 19 22:22:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	100180	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.710	114	215585	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	201304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.447	152	71991	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	45263	45.714	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.420%
35) Dibromofluoromethane	7.740	113	50517	36.772	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	73.540%
50) Toluene-d8	10.197	98	187876	39.715	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	79.420%
62) 4-Bromofluorobenzene	12.502	95	55507	34.777	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	69.560%
Target Compounds						
20) Methylene Chloride	4.747	84	19480	9.170	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011924\
Data File : VY017054.D
Acq On : 19 Jan 2024 14:35
Operator : SY/MD
Sample : P1185-01
Misc : 3.44g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-7-19.5-20.0

Quant Time: Jan 19 22:22:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

