

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011924\
 Data File : VY017056.D
 Acq On : 19 Jan 2024 15:22
 Operator : SY/MD
 Sample : P1185-03
 Misc : 4.17g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-6-10.5-11.0

Quant Time: Jan 19 22:23:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	95161	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.715	114	205128	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.514	117	196311	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	72599	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	44793	47.625	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.260%
35) Dibromofluoromethane	7.746	113	48570	37.157	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	74.320%
50) Toluene-d8	10.203	98	182183	40.474	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	80.940%
62) 4-Bromofluorobenzene	12.501	95	55832	36.764	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	73.520%
Target Compounds						
20) Methylene Chloride	4.759	84	16980	7.769	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011924\
Data File : VY017056.D
Acq On : 19 Jan 2024 15:22
Operator : SY/MD
Sample : P1185-03
Misc : 4.17g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-6-10.5-11.0

Quant Time: Jan 19 22:23:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

