

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY012020\  
 Data File : VY001293.D  
 Acq On : 20 Jan 2020 14:20  
 Operator : SY/MD  
 Sample : VY0120SBSD01  
 Misc : 5.00G/5ML/MSVOA Y/SOIL  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampled :

Quant Time: Jan 21 07:01:00 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_Y\METHODS\82Y011020S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jan 21 06:54:04 2020  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 7.82  | 168  | 192924   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 8.71  | 114  | 336295   | 50.00 | ug/l  | -0.01    |
| 63) Chlorobenzene-d5       | 11.50 | 117  | 303502   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.44 | 152  | 145527   | 50.00 | ug/l  | 0.00     |

## System Monitoring Compounds

|                           |        |     |          |           |      |       |
|---------------------------|--------|-----|----------|-----------|------|-------|
| 33) 1,2-Dichloroethane-d4 | 8.16   | 65  | 102750   | 50.43     | ug/l | -0.01 |
| Spiked Amount             | 50.000 |     | Recovery | = 100.86% |      |       |
| 35) Dibromofluoromethane  | 7.74   | 113 | 94745    | 48.12     | ug/l | -0.01 |
| Spiked Amount             | 50.000 |     | Recovery | = 96.24%  |      |       |
| 50) Toluene-d8            | 10.19  | 98  | 393481   | 51.86     | ug/l | 0.00  |
| Spiked Amount             | 50.000 |     | Recovery | = 103.72% |      |       |
| 62) 4-Bromofluorobenzene  | 12.49  | 95  | 144070   | 46.45     | ug/l | 0.00  |
| Spiked Amount             | 50.000 |     | Recovery | = 92.90%  |      |       |

## Target Compounds

|                               |      |     |        |         |      | Qvalue |
|-------------------------------|------|-----|--------|---------|------|--------|
| 2) Dichlorodifluoromethane    | 1.91 | 85  | 27287  | 18.417  | ug/l | 96     |
| 3) Chloromethane              | 2.12 | 50  | 43424  | 20.333  | ug/l | 99     |
| 4) Vinyl Chloride             | 2.26 | 62  | 45797  | 21.286  | ug/l | 97     |
| 5) Bromomethane               | 2.65 | 94  | 30867  | 26.708  | ug/l | 96     |
| 6) Chloroethane               | 2.80 | 64  | 29580  | 23.797  | ug/l | 98     |
| 7) Trichlorofluoromethane     | 3.13 | 101 | 60084  | 20.063  | ug/l | 94     |
| 8) Diethyl Ether              | 3.54 | 74  | 25414  | 21.839  | ug/l | 96     |
| 9) 1,1,2-Trichlorotrifluoroet | 3.91 | 101 | 38212  | 19.822  | ug/l | 99     |
| 10) Methyl Iodide             | 4.11 | 142 | 49080  | 18.601  | ug/l | 97     |
| 11) Tert butyl alcohol        | 4.98 | 59  | 33612  | 163.952 | ug/l | # 82   |
| 12) 1,1-Dichloroethene        | 3.89 | 96  | 39635  | 20.096  | ug/l | 98     |
| 13) Acrolein                  | 3.74 | 56  | 21284  | 115.322 | ug/l | 94     |
| 14) Allyl chloride            | 4.50 | 41  | 70714  | 21.244  | ug/l | 99     |
| 15) Acrylonitrile             | 5.19 | 53  | 65073  | 111.809 | ug/l | 99     |
| 16) Acetone                   | 3.97 | 43  | 57550  | 121.614 | ug/l | 99     |
| 17) Carbon Disulfide          | 4.21 | 76  | 119693 | 18.751  | ug/l | 100    |
| 18) Methyl Acetate            | 4.50 | 43  | 34056  | 21.767  | ug/l | 96     |
| 19) Methyl tert-butyl Ether   | 5.24 | 73  | 117479 | 21.534  | ug/l | 98     |
| 20) Methylene Chloride        | 4.74 | 84  | 56990  | 22.842  | ug/l | 93     |
| 21) trans-1,2-Dichloroethene  | 5.24 | 96  | 46224  | 20.670  | ug/l | 94     |
| 22) Diisopropyl ether         | 6.15 | 45  | 149700 | 22.019  | ug/l | 99     |
| 23) Vinyl Acetate             | 6.08 | 43  | 482196 | 110.899 | ug/l | 99     |
| 24) 1,1-Dichloroethane        | 6.05 | 63  | 80715  | 21.686  | ug/l | 99     |
| 25) 2-Butanone                | 7.01 | 43  | 89321  | 121.278 | ug/l | 97     |
| 26) 2,2-Dichloropropane       | 7.01 | 77  | 70564  | 20.617  | ug/l | 98     |
| 27) cis-1,2-Dichloroethene    | 7.01 | 96  | 53544  | 21.851  | ug/l | 99     |
| 28) Bromochloromethane        | 7.35 | 49  | 30714  | 20.095  | ug/l | 95     |
| 29) Tetrahydrofuran           | 7.37 | 42  | 56539  | 112.020 | ug/l | 96     |
| 30) Chloroform                | 7.53 | 83  | 78021  | 21.394  | ug/l | 99     |
| 31) Cyclohexane               | 7.80 | 56  | 75817  | 19.239  | ug/l | 96     |
| 32) 1,1,1-Trichloroethane     | 7.72 | 97  | 66159  | 20.446  | ug/l | 99     |
| 36) 1,1-Dichloropropene       | 7.94 | 75  | 62307  | 20.328  | ug/l | 99     |
| 37) Ethyl Acetate             | 7.10 | 43  | 39680  | 22.059  | ug/l | 96     |
| 38) Carbon Tetrachloride      | 7.92 | 117 | 56455  | 20.098  | ug/l | 94     |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY012020\  
 Data File : VY001293.D  
 Acq On : 20 Jan 2020 14:20  
 Operator : SY/MD  
 Sample : VY0120SBSD01  
 Misc : 5.00G/5ML/MSVOA Y/SOIL  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :

Quant Time: Jan 21 07:01:00 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_Y\METHODS\82Y011020S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jan 21 06:54:04 2020  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc    | Units  | Dev(Min) |
|--------------------------------|-------|------|----------|---------|--------|----------|
| 39) Methylcyclohexane          | 9.20  | 83   | 79409    | 19.110  | ug/l   | 99       |
| 40) Benzene                    | 8.18  | 78   | 190686   | 21.312  | ug/l   | 100      |
| 41) Methacrylonitrile          | 7.36  | 41   | 51724    | 61.213  | ug/l # | 1        |
| 42) 1,2-Dichloroethane         | 8.26  | 62   | 53824    | 21.484  | ug/l   | 98       |
| 43) Isopropyl Acetate          | 8.29  | 43   | 74702    | 21.717  | ug/l   | 98       |
| 44) Trichloroethene            | 8.96  | 130  | 49464    | 21.216  | ug/l   | 99       |
| 45) 1,2-Dichloropropane        | 9.24  | 63   | 47897    | 21.654  | ug/l   | 99       |
| 46) Dibromomethane             | 9.32  | 93   | 25668    | 21.268  | ug/l   | 99       |
| 47) Bromodichloromethane       | 9.51  | 83   | 62415    | 21.220  | ug/l   | 99       |
| 48) Methyl methacrylate        | 9.31  | 41   | 33944    | 21.325  | ug/l   | 99       |
| 49) 1,4-Dioxane                | 9.31  | 88   | 7178     | 488.378 | ug/l # | 91       |
| 51) 4-Methyl-2-Pentanone       | 10.08 | 43   | 192002   | 110.007 | ug/l   | 98       |
| 52) Toluene                    | 10.25 | 92   | 122263   | 21.425  | ug/l   | 98       |
| 53) t-1,3-Dichloropropene      | 10.48 | 75   | 68144    | 20.704  | ug/l   | 97       |
| 54) cis-1,3-Dichloropropene    | 9.94  | 75   | 78643    | 21.193  | ug/l   | 98       |
| 55) 1,1,2-Trichloroethane      | 10.66 | 97   | 39035    | 21.817  | ug/l   | 97       |
| 56) Ethyl methacrylate         | 10.52 | 69   | 57605    | 21.048  | ug/l   | 100      |
| 57) 1,3-Dichloropropane        | 10.80 | 76   | 65699    | 21.389  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether  | 9.80  | 63   | 100635   | 104.546 | ug/l   | 100      |
| 59) 2-Hexanone                 | 10.85 | 43   | 134327   | 109.815 | ug/l   | 98       |
| 60) Dibromochloromethane       | 11.00 | 129  | 43927    | 21.324  | ug/l   | 99       |
| 61) 1,2-Dibromoethane          | 11.10 | 107  | 36030    | 20.624  | ug/l   | 98       |
| 64) Tetrachloroethene          | 10.73 | 164  | 37713    | 19.276  | ug/l   | 97       |
| 65) Chlorobenzene              | 11.53 | 112  | 126438   | 20.849  | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane  | 11.60 | 131  | 44868    | 21.466  | ug/l   | 99       |
| 67) Ethyl Benzene              | 11.60 | 91   | 232867   | 20.991  | ug/l   | 99       |
| 68) m/p-Xylenes                | 11.71 | 106  | 174868   | 41.785  | ug/l   | 100      |
| 69) o-Xylene                   | 12.04 | 106  | 84683    | 21.105  | ug/l   | 99       |
| 70) Styrene                    | 12.05 | 104  | 146886   | 20.864  | ug/l   | 99       |
| 71) Bromoform                  | 12.22 | 173  | 25689    | 19.929  | ug/l # | 100      |
| 73) Isopropylbenzene           | 12.34 | 105  | 225149   | 20.580  | ug/l   | 100      |
| 74) N-amyl acetate             | 12.16 | 43   | 70523    | 20.655  | ug/l   | 100      |
| 75) 1,1,2,2-Tetrachloroethane  | 12.60 | 83   | 47283    | 20.542  | ug/l   | 100      |
| 76) 1,2,3-Trichloropropane     | 12.64 | 75   | 61910    | 39.740  | ug/l # | 100      |
| 77) Bromobenzene               | 12.62 | 156  | 50860    | 21.039  | ug/l   | 95       |
| 78) n-propylbenzene            | 12.68 | 91   | 268341   | 20.391  | ug/l   | 99       |
| 79) 2-Chlorotoluene            | 12.77 | 91   | 155107   | 20.523  | ug/l   | 100      |
| 80) 1,3,5-Trimethylbenzene     | 12.82 | 105  | 188283   | 20.626  | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-buten | 12.39 | 75   | 17164    | 18.683  | ug/l # | 96       |
| 82) 4-Chlorotoluene            | 12.86 | 91   | 163174   | 20.224  | ug/l   | 99       |
| 83) tert-Butylbenzene          | 13.08 | 119  | 155609   | 20.287  | ug/l   | 99       |
| 84) 1,2,4-Trimethylbenzene     | 13.13 | 105  | 189934   | 20.952  | ug/l   | 99       |
| 85) sec-Butylbenzene           | 13.27 | 105  | 226969   | 20.569  | ug/l   | 99       |
| 86) p-Isopropyltoluene         | 13.38 | 119  | 202499   | 21.136  | ug/l   | 99       |
| 87) 1,3-Dichlorobenzene        | 13.38 | 146  | 99646    | 21.796  | ug/l   | 98       |
| 88) 1,4-Dichlorobenzene        | 13.45 | 146  | 99963    | 21.245  | ug/l   | 99       |
| 89) n-Butylbenzene             | 13.70 | 91   | 204497   | 21.193  | ug/l   | 98       |
| 90) Hexachloroethane           | 13.97 | 117  | 38021    | 20.310  | ug/l   | 99       |
| 91) 1,2-Dichlorobenzene        | 13.75 | 146  | 90756    | 21.360  | ug/l   | 98       |
| 92) 1,2-Dibromo-3-Chloropropan | 14.36 | 75   | 8293     | 19.973  | ug/l   | 98       |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY012020\  
 Data File : VY001293.D  
 Acq On : 20 Jan 2020 14:20  
 Operator : SY/MD  
 Sample : VY0120SBSD01  
 Misc : 5.00G/5ML/MSVOA Y/SOIL  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :

Quant Time: Jan 21 07:01:00 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_Y\METHODS\82Y011020S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jan 21 06:54:04 2020  
 Response via : Initial Calibration


| Internal Standards         | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|-------|------|----------|--------|-------|----------|
| 93) 1,2,4-Trichlorobenzene | 15.02 | 180  | 62477    | 22.317 | ug/l  | 99       |
| 94) Hexachlorobutadiene    | 15.12 | 225  | 30315    | 21.495 | ug/l  | 99       |
| 95) Naphthalene            | 15.25 | 128  | 138436   | 21.241 | ug/l  | 99       |
| 96) 1,2,3-Trichlorobenzene | 15.43 | 180  | 54700    | 22.171 | ug/l  | 99       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY012020\  
 Data File : VY001293.D  
 Acq On : 20 Jan 2020 14:20  
 Operator : SY/MD  
 Sample : VY0120SBSD01  
 Misc : 5.00G/5ML/MSVOA Y/SOIL  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 Client Sampled :

Quant Time: Jan 21 07:01:00 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_Y\METHODS\82Y011020S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jan 21 06:54:04 2020  
 Response via : Initial Calibration









#1  
 Pentafluorobenzene  
 Concen: 50.000 ug/l  
 RT: 7.82 min Scan# 997  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 168     | 100  |       |       |
| 99      | 55.5 | 47.9  | 71.9  |



#2  
 Dichlorodifluoromethane  
 Concen: 18.417 ug/l  
 RT: 1.91 min Scan# 28  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 85      | 100  |       |       |
| 87      | 30.3 | 16.4  | 49.2  |









#3  
 Chloromethane  
 Concen: 20.333 ug/l  
 RT: 2.12 min Scan# 63  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 50      | 43424 |       |       |
| 52      | 31.5  | 25.8  | 38.6  |



#4  
 Vinyl Chloride  
 Concen: 21.286 ug/l  
 RT: 2.26 min Scan# 86  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 62      | 45797 |       |       |
| 64      | 30.5  | 25.8  | 38.8  |








#5  
 Bromomethane  
 Concen: 26.708 ug/l  
 RT: 2.65 min Scan# 150  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 94      | 30867 |       |       |
| 96      | 92.5  | 77.3  | 115.9 |



#6  
 Chloroethane  
 Concen: 23.797 ug/l  
 RT: 2.80 min Scan# 175  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 64      | 29580 |       |       |
| 66      | 33.0  | 25.5  | 38.3  |









#7  
 Trichlorofluoromethane  
 Concen: 20.063 ug/l  
 RT: 3.13 min Scan# 229  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :

Tgt Ion:101 Resp: 60084  
 Ion Ratio Lower Upper  
 101 100  
 103 61.4 52.6 79.0




#8  
 Diethyl Ether  
 Concen: 21.839 ug/l  
 RT: 3.54 min Scan# 296  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

Tgt Ion: 74 Resp: 25414  
 Ion Ratio Lower Upper  
 74 100  
 45 94.4 45.5 136.5











#9  
 1,1,2-Trichlorotrifluoroethane  
 Concen: 19.822 ug/l  
 RT: 3.91 min Scan# 357  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 101     | 38212 |       |       |
| 101     | 100   |       |       |
| 85      | 44.6  | 35.2  | 52.8  |
| 151     | 76.4  | 61.4  | 92.2  |



#10  
 Methyl Iodide  
 Concen: 18.601 ug/l  
 RT: 4.11 min Scan# 389  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 142     | 49080 |       |       |
| 142     | 100   |       |       |
| 127     | 40.4  | 30.8  | 46.2  |
| 141     | 14.3  | 11.0  | 16.4  |









#11  
 Tert butyl alcohol  
 Concen: 163.952 ug/l  
 RT: 4.98 min Scan# 532  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 59      | 100  |       |       |
| 57      | 3.3  | 7.8   | 11.8# |



#12  
 1,1-Dichloroethene  
 Concen: 20.096 ug/l  
 RT: 3.89 min Scan# 353  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 96      | 100   |       |       |
| 61      | 154.5 | 122.1 | 183.1 |
| 98      | 62.0  | 51.0  | 76.6  |









#13  
 Acrolein  
 Concen: 115.322 ug/l  
 RT: 3.74 min Scan# 329  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 56      | 21284 |       |       |
| 55      | 76.2  | 56.7  | 85.1  |



#14  
 Allyl chloride  
 Concen: 21.244 ug/l  
 RT: 4.50 min Scan# 453  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 41      | 70714 |       |       |
| 39      | 64.9  | 52.6  | 78.8  |
| 76      | 37.1  | 30.2  | 45.2  |









#15  
 Acrylonitrile  
 Concen: 111.809 ug/l  
 RT: 5.19 min Scan# 566  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 53      | 100  |       |       |
| 52      | 80.8 | 65.7  | 98.5  |
| 51      | 35.6 | 28.7  | 43.1  |



#16  
 Acetone  
 Concen: 121.614 ug/l  
 RT: 3.97 min Scan# 366  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 43      | 100  |       |       |
| 58      | 31.8 | 25.8  | 38.6  |









#17  
 Carbon Disulfide  
 Concen: 18.751 ug/l  
 RT: 4.21 min Scan# 405  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 76      | 119693 |       |       |
| 78      | 8.9    | 7.2   | 10.8  |



#18  
 Methyl Acetate  
 Concen: 21.767 ug/l  
 RT: 4.50 min Scan# 453  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 43      | 34056 |       |       |
| 74      | 23.4  | 20.2  | 30.2  |









#19  
 Methyl tert-butyl Ether  
 Concen: 21.534 ug/l  
 RT: 5.24 min Scan# 575  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 73      | 117479 |       |       |
| 57      | 21.0   | 17.8  | 26.6  |



#20  
 Methylene Chloride  
 Concen: 22.842 ug/l  
 RT: 4.74 min Scan# 493  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 84      | 56990 |       |       |
| 49      | 129.4 | 93.8  | 140.6 |
| 51      | 39.7  | 29.9  | 44.9  |
| 86      | 67.1  | 52.6  | 78.8  |









#21  
 trans-1,2-Dichloroethene  
 Concen: 20.670 ug/l  
 RT: 5.24 min Scan# 575  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 96      | 46224 |       |       |
| 61      | 127.9 | 107.6 | 161.4 |
| 98      | 65.4  | 47.6  | 71.4  |



#22  
 Diisopropyl ether  
 Concen: 22.019 ug/l  
 RT: 6.15 min Scan# 723  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 45      | 149700 |       |       |
| 43      | 56.4   | 46.5  | 69.7  |
| 87      | 28.1   | 22.5  | 33.7  |
| 59      | 12.1   | 9.4   | 14.0  |









#23  
 Vinyl Acetate  
 Concen: 110.899 ug/l  
 RT: 6.08 min Scan# 713  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 43      | 100  |       |       |
| 86      | 10.7 | 9.0   | 13.4  |




#24  
 1,1-Dichloroethane  
 Concen: 21.686 ug/l  
 RT: 6.05 min Scan# 707  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 63      | 100  |       |       |
| 98      | 7.3  | 3.9   | 11.7  |
| 100     | 5.3  | 2.4   | 7.2   |











#25  
 2-Butanone  
 Concen: 121.278 ug/l  
 RT: 7.01 min Scan# 865  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


Tgt Ion: 43 Resp: 89321  
 Ion Ratio Lower Upper  
 43 100  
 72 25.9 22.2 33.2



#26  
 2,2-Dichloropropane  
 Concen: 20.617 ug/l  
 RT: 7.01 min Scan# 865  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

Tgt Ion: 77 Resp: 70564  
 Ion Ratio Lower Upper  
 77 100  
 97 22.7 11.9 35.7









#27  
 cis-1,2-Dichloroethene  
 Concen: 21.851 ug/l  
 RT: 7.01 min Scan# 865  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 96      | 53544 |       |       |
| 96      | 100   |       |       |
| 61      | 143.1 | 0.0   | 284.0 |
| 98      | 65.2  | 0.0   | 130.2 |



#28  
 Bromochloromethane  
 Concen: 20.095 ug/l  
 RT: 7.35 min Scan# 921  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 49      | 30714 |       |       |
| 49      | 100   |       |       |
| 129     | 1.3   | 0.0   | 4.4   |
| 130     | 72.4  | 61.2  | 91.8  |









#29  
 Tetrahydrofuran  
 Concen: 112.020 ug/l  
 RT: 7.37 min Scan# 924  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 42      | 56539 |       |       |
| 72      | 43.0  | 36.9  | 55.3  |
| 71      | 40.7  | 34.0  | 51.0  |



#30  
 Chloroform  
 Concen: 21.394 ug/l  
 RT: 7.53 min Scan# 950  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 83      | 78021 |       |       |
| 85      | 66.0  | 51.9  | 77.9  |









#31  
 Cyclohexane  
 Concen: 19.239 ug/l  
 RT: 7.80 min Scan# 995  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 56      | 100  |       |       |
| 69      | 39.1 | 27.8  | 41.8  |
| 84      | 88.4 | 68.9  | 103.3 |



#32  
 1,1,1-Trichloroethane  
 Concen: 20.446 ug/l  
 RT: 7.72 min Scan# 982  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 97      | 100  |       |       |
| 99      | 63.7 | 51.3  | 76.9  |
| 61      | 43.9 | 34.6  | 52.0  |









#33  
 1,2-Dichloroethane-d4  
 Concen: 50.427 ug/l  
 RT: 8.16 min Scan# 1054  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 65      | 102750 |       |       |
| 65      | 100    |       |       |
| 67      | 54.4   | 0.0   | 105.2 |



#34  
 1,4-Difluorobenzene  
 Concen: 50.000 ug/l  
 RT: 8.71 min Scan# 1143  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 114     | 336295 |       |       |
| 114     | 100    |       |       |
| 63      | 20.9   | 0.0   | 40.8  |
| 88      | 16.5   | 0.0   | 32.4  |









#35  
 Dibromofluoromethane  
 Concen: 48.118 ug/l  
 RT: 7.74 min Scan# 985  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 113     | 100   |       |       |
| 111     | 103.6 | 82.9  | 124.3 |
| 192     | 19.0  | 15.4  | 23.2  |



#36  
 1,1-Dichloropropene  
 Concen: 20.328 ug/l  
 RT: 7.94 min Scan# 1017  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 75      | 100  |       |       |
| 110     | 37.0 | 18.3  | 54.8  |
| 77      | 31.3 | 25.1  | 37.7  |








#37  
 Ethyl Acetate  
 Concen: 22.059 ug/l  
 RT: 7.10 min Scan# 880  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20





Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 43      | 100  |       |       |
| 61      | 12.5 | 11.1  | 16.7  |
| 70      | 10.5 | 9.7   | 14.5  |



#38  
 Carbon Tetrachloride  
 Concen: 20.098 ug/l  
 RT: 7.92 min Scan# 1014  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 117     | 100  |       |       |
| 119     | 91.1 | 78.3  | 117.5 |
| 121     | 30.4 | 23.5  | 35.3  |









#39  
 Methylcyclohexane  
 Concen: 19.110 ug/l  
 RT: 9.20 min Scan# 1224  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 83      | 100  |       |       |
| 55      | 74.6 | 58.9  | 88.3  |
| 98      | 49.3 | 38.4  | 57.6  |




#40  
 Benzene  
 Concen: 21.312 ug/l  
 RT: 8.18 min Scan# 1057  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 78      | 100  |       |       |
| 77      | 23.5 | 18.8  | 28.2  |











#41  
 Methacrylonitrile  
 Concen: 61.213 ug/l  
 RT: 7.36 min Scan# 923  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

Instrument :  
 MSVOA\_Y  
 ClientSampled :

| Tgt Ion | Resp | Lower | Upper  |
|---------|------|-------|--------|
| 41      | 100  |       |        |
| 39      | 0.0  | 103.5 | 155.3# |
| 67      | 0.0  | 0.0   | 0.0    |
| 52      | 0.0  | 0.0   | 0.0    |




Abundance Ion 41.00 (40.70 to 41.70): VY0  
 Ion 39.00 (38.70 to 39.70): VY0  
 Ion 67.00 (66.70 to 67.70): VY0  
 Ion 52.00 (51.70 to 52.70): VY0






#42  
 1,2-Dichloroethane  
 Concen: 21.484 ug/l  
 RT: 8.26 min Scan# 1070  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 62      | 100  |       |       |
| 98      | 10.8 | 0.0   | 20.4  |



Abundance Ion 61.90 (61.60 to 62.60): VY0  
 Ion 97.90 (97.60 to 98.60): VY0









#43

Isopropyl Acetate  
 Concen: 21.717 ug/l  
 RT: 8.29 min Scan# 1075  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

Instrument :  
 MSVOA\_Y  
 ClientSampled :




| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 43      | 100  |       |       |
| 61      | 28.8 | 23.9  | 35.9  |
| 87      | 13.3 | 11.4  | 17.2  |




#44

Trichloroethene  
 Concen: 21.216 ug/l  
 RT: 8.96 min Scan# 1184  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 130     | 100  |       |       |
| 95      | 97.7 | 0.0   | 197.6 |









#45  
 1,2-Dichloropropane  
 Concen: 21.654 ug/l  
 RT: 9.24 min Scan# 1230  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 63      | 47897 |       |       |
| 65      | 32.9  | 25.8  | 38.8  |



#46  
 Dibromomethane  
 Concen: 21.268 ug/l  
 RT: 9.32 min Scan# 1244  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 93      | 25668 |       |       |
| 95      | 81.3  | 66.8  | 100.2 |
| 174     | 97.3  | 78.4  | 117.6 |








#47

Bromodichloromethane  
 Concen: 21.220 ug/l  
 RT: 9.51 min Scan# 1275  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

Instrument :  
 MSVOA\_Y  
 ClientSampled :





| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 83      | 62415 |       |       |
| 85      | 63.5  | 51.5  | 77.3  |
| 127     | 7.9   | 6.3   | 9.5   |




#48

Methyl methacrylate  
 Concen: 21.325 ug/l  
 RT: 9.31 min Scan# 1242  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 41      | 33944 |       |       |
| 69      | 88.1  | 69.8  | 104.8 |
| 39      | 56.3  | 45.2  | 67.8  |









#49  
 1,4-Dioxane  
 Concen: 488.378 ug/l  
 RT: 9.31 min Scan# 1242  
 Delta R.T. -0.02 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 88      | 7178 |       |       |
| 88      | 100  |       |       |
| 43      | 37.6 | 22.4  | 33.6# |
| 58      | 77.1 | 58.0  | 87.0  |



#50  
 Toluene-d8  
 Concen: 51.863 ug/l  
 RT: 10.19 min Scan# 1387  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 98      | 393481 |       |       |
| 98      | 100    |       |       |
| 100     | 65.5   | 53.1  | 79.7  |









#51  
 4-Methyl-2-Pentanone  
 Concen: 110.007 ug/l  
 RT: 10.08 min Scan# 1369  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 43      | 192002 |       |       |
| 58      | 39.4   | 32.3  | 48.5  |



#52  
 Toluene  
 Concen: 21.425 ug/l  
 RT: 10.25 min Scan# 1397  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 92      | 122263 |       |       |
| 91      | 168.1  | 136.6 | 204.8 |









#53  
 t-1,3-Dichloropropene  
 Concen: 20.704 ug/l  
 RT: 10.48 min Scan# 1434  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 75      | 68144 |       |       |
| 77      | 31.1  | 26.2  | 39.2  |



#54  
 cis-1,3-Dichloropropene  
 Concen: 21.193 ug/l  
 RT: 9.94 min Scan# 1346  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 75      | 78643 |       |       |
| 77      | 31.5  | 25.3  | 37.9  |
| 39      | 46.5  | 39.1  | 58.7  |









#55  
 1,1,2-Trichloroethane  
 Concen: 21.817 ug/l  
 RT: 10.66 min Scan# 1463  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 97      | 39035 |       |       |
| 97      | 100   |       |       |
| 83      | 80.6  | 67.2  | 100.8 |
| 85      | 56.4  | 44.9  | 67.3  |
| 99      | 64.8  | 49.9  | 74.9  |




#56  
 Ethyl methacrylate  
 Concen: 21.048 ug/l  
 RT: 10.52 min Scan# 1441  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 69      | 57605 |       |       |
| 69      | 100   |       |       |
| 41      | 64.2  | 51.4  | 77.2  |
| 39      | 32.3  | 26.3  | 39.5  |










#57  
 1,3-Dichloropropane  
 Concen: 21.389 ug/l  
 RT: 10.80 min Scan# 1487  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20





Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 76      | 100  |       |       |
| 78      | 32.2 | 26.1  | 39.1  |



#58  
 2-Chloroethyl Vinyl ether  
 Concen: 104.546 ug/l  
 RT: 9.80 min Scan# 1322  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 63      | 100  |       |       |
| 106     | 28.8 | 23.0  | 34.6  |









#59  
 2-Hexanone  
 Concen: 109.815 ug/l  
 RT: 10.85 min Scan# 1494  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


Tgt Ion: 43 Resp: 134327  
 Ion Ratio Lower Upper  
 43 100  
 58 55.0 28.4 85.2



#60  
 Dibromochloromethane  
 Concen: 21.324 ug/l  
 RT: 11.00 min Scan# 1519  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

Tgt Ion: 129 Resp: 43927  
 Ion Ratio Lower Upper  
 129 100  
 127 77.1 38.9 116.7









#61  
 1,2-Dibromoethane  
 Concen: 20.624 ug/l  
 RT: 11.10 min Scan# 1536  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


Tgt Ion: 107 Resp: 36030  
 Ion Ratio Lower Upper  
 107 100  
 109 95.8 75.4 113.0



#62  
 4-Bromofluorobenzene  
 Concen: 46.448 ug/l  
 RT: 12.49 min Scan# 1764  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

Tgt Ion: 95 Resp: 144070  
 Ion Ratio Lower Upper  
 95 100  
 174 74.4 0.0 143.2  
 176 71.7 0.0 136.0








#63  
 Chlorobenzene-d5  
 Concen: 50.000 ug/l  
 RT: 11.50 min Scan# 1602  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20





Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 117     | 303502 |       |       |
| 82      | 53.1   | 44.9  | 67.3  |
| 119     | 32.1   | 25.7  | 38.5  |



#64  
 Tetrachloroethene  
 Concen: 19.276 ug/l  
 RT: 10.73 min Scan# 1475  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 164     | 37713 |       |       |
| 166     | 132.1 | 102.3 | 153.5 |
| 129     | 96.2  | 76.2  | 114.4 |
| 131     | 97.2  | 73.9  | 110.9 |









#65  
 Chlorobenzene  
 Concen: 20.849 ug/l  
 RT: 11.53 min Scan# 1606  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 112     | 126438 |       |       |
| 114     | 33.0   | 25.9  | 38.9  |



#66  
 1,1,1,2-Tetrachloroethane  
 Concen: 21.466 ug/l  
 RT: 11.60 min Scan# 1618  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 131     | 44868 |       |       |
| 133     | 93.6  | 47.3  | 142.0 |
| 119     | 65.8  | 33.6  | 100.8 |








#67  
Ethyl Benzene  
Concen: 20.991 ug/l  
RT: 11.60 min Scan# 1618  
Delta R.T. -0.01 min  
Lab File: VY001293.D  
Acq: 20 Jan 2020 14:20




Instrument : MSVOA\_Y  
Client Sampled :


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 91    | Resp: | 232867 |
| Ion Ratio | Lower | Upper |        |
| 91        | 100   |       |        |
| 106       | 31.3  | 25.3  | 37.9   |



#68  
m/p-Xylenes  
Concen: 41.785 ug/l  
RT: 11.71 min Scan# 1636  
Delta R.T. -0.01 min  
Lab File: VY001293.D  
Acq: 20 Jan 2020 14:20

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 106   | Resp: | 174868 |
| Ion Ratio | Lower | Upper |        |
| 106       | 100   |       |        |
| 91        | 201.3 | 161.4 | 242.2  |









#69  
 o-Xylene  
 Concen: 21.105 ug/l  
 RT: 12.04 min Scan# 1690  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument : MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 106     | 84683 |       |       |
| 91      | 210.9 | 106.3 | 318.8 |



#70  
 Styrene  
 Concen: 20.864 ug/l  
 RT: 12.05 min Scan# 1692  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 104     | 146886 |       |       |
| 78      | 48.9   | 40.5  | 60.7  |
| 103     | 52.8   | 42.4  | 63.6  |









#71  
 Bromoform  
 Concen: 19.929 ug/l  
 RT: 12.22 min Scan# 1719  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 173     | 100  |       |       |
| 175     | 48.4 | 24.1  | 72.2  |
| 254     | 0.0  | 0.1   | 0.1#  |




#72  
 1,4-Dichlorobenzene-d4  
 Concen: 50.000 ug/l  
 RT: 13.44 min Scan# 1919  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 152     | 100   |       |       |
| 115     | 60.3  | 30.9  | 92.5  |
| 150     | 163.5 | 0.0   | 346.8 |











#73  
 Isopropylbenzene  
 Concen: 20.580 ug/l  
 RT: 12.34 min Scan# 1739  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument : MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 105     | 100  |       |       |
| 120     | 26.6 | 13.4  | 40.2  |



#74  
 N-nyl acetate  
 Concen: 20.655 ug/l  
 RT: 12.16 min Scan# 1709  
 Delta R.T. 0.00 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 43      | 100  |       |       |
| 70      | 44.1 | 35.0  | 52.6  |
| 55      | 22.8 | 18.4  | 27.6  |
| 61      | 24.9 | 20.1  | 30.1  |









#75  
 1,1,2,2-Tetrachloroethane  
 Concen: 20.542 ug/l  
 RT: 12.60 min Scan# 1781  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 83      | 47283 |       |       |
| 131     | 10.6  | 5.1   | 15.3  |
| 85      | 64.7  | 32.2  | 96.6  |



#76  
 1,2,3-Trichloropropane  
 Concen: 39.740 ug/l  
 RT: 12.64 min Scan# 1789  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 75      | 61910 |       |       |
| 77      | 0.0   | 0.0   | 0.0   |









#77  
 Bromobenzene  
 Concen: 21.039 ug/l  
 RT: 12.62 min Scan# 1785  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 156     | 100   |       |       |
| 77      | 202.4 | 106.7 | 320.1 |
| 158     | 95.2  | 48.3  | 144.8 |



#78  
 n-propylbenzene  
 Concen: 20.391 ug/l  
 RT: 12.68 min Scan# 1795  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 91      | 100  |       |       |
| 120     | 23.1 | 11.5  | 34.4  |









#79  
 2-Chlorotoluene  
 Concen: 20.523 ug/l  
 RT: 12.77 min Scan# 1809  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument : MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 91      | 155107 |       |       |
| 126     | 32.8   | 16.4  | 49.4  |



#80  
 1,3,5-Trimethylbenzene  
 Concen: 20.626 ug/l  
 RT: 12.82 min Scan# 1818  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 105     | 188283 |       |       |
| 120     | 49.3   | 24.2  | 72.6  |









#81  
 trans-1,4-Dichloro-2-butene  
 Concen: 18.683 ug/l  
 RT: 12.39 min Scan# 1747  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 75      | 17164 |       |       |
| 75      | 100   |       |       |
| 53      | 90.3  | 69.4  | 104.2 |
| 89      | 45.8  | 0.0   | 0.0#  |



#82  
 4-Chlorotoluene  
 Concen: 20.224 ug/l  
 RT: 12.86 min Scan# 1825  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 91      | 163174 |       |       |
| 91      | 100    |       |       |
| 126     | 33.0   | 16.3  | 48.8  |









#83  
 tert-Butylbenzene  
 Concen: 20.287 ug/l  
 RT: 13.08 min Scan# 1861  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 119     | 155609 |       |       |
| 91      | 68.1   | 33.9  | 101.7 |
| 134     | 27.6   | 13.5  | 40.4  |



#84  
 1,2,4-Trimethylbenzene  
 Concen: 20.952 ug/l  
 RT: 13.13 min Scan# 1869  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 105     | 189934 |       |       |
| 120     | 46.1   | 22.8  | 68.3  |









#85  
 sec-Butylbenzene  
 Concen: 20.569 ug/l  
 RT: 13.27 min Scan# 1891  
 Delta R.T. 0.00 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 105     | 100  |       |       |
| 134     | 20.3 | 9.8   | 29.5  |



#86  
 p-Isopropyltoluene  
 Concen: 21.136 ug/l  
 RT: 13.38 min Scan# 1910  
 Delta R.T. 0.00 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 119     | 100  |       |       |
| 134     | 27.1 | 13.4  | 40.2  |
| 91      | 24.2 | 12.4  | 37.2  |









#87  
 1,3-Dichlorobenzene  
 Concen: 21.796 ug/l  
 RT: 13.38 min Scan# 1909  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 ClientSampled :

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 99646 |       |       |
| 146     | 100   |       |       |
| 111     | 40.1  | 21.3  | 64.0  |
| 148     | 63.9  | 31.9  | 95.5  |




#88  
 1,4-Dichlorobenzene  
 Concen: 21.245 ug/l  
 RT: 13.45 min Scan# 1922  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 99963 |       |       |
| 146     | 100   |       |       |
| 111     | 40.9  | 20.8  | 62.5  |
| 148     | 63.3  | 32.0  | 96.2  |











#89  
 n-Butylbenzene  
 Concen: 21.193 ug/l  
 RT: 13.70 min Scan# 1963  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument :  
 MSVOA\_Y  
 Client Sampled :


| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 91      | 100  |       |       |
| 92      | 53.9 | 27.7  | 83.1  |
| 134     | 25.5 | 12.9  | 38.6  |



#90  
 Hexachloroethane  
 Concen: 20.310 ug/l  
 RT: 13.97 min Scan# 2007  
 Delta R.T. 0.00 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 117     | 100  |       |       |
| 201     | 71.4 | 35.4  | 106.2 |









#91  
 1,2-Dichlorobenzene  
 Concen: 21.360 ug/l  
 RT: 13.75 min Scan# 1970  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument : MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 146     | 100  |       |       |
| 111     | 42.0 | 22.0  | 66.0  |
| 148     | 62.9 | 32.1  | 96.3  |



#92  
 1,2-Dibromo-3-Chloropropane  
 Concen: 19.973 ug/l  
 RT: 14.36 min Scan# 2071  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 75      | 100   |       |       |
| 155     | 85.1  | 41.2  | 123.6 |
| 157     | 107.3 | 52.8  | 158.4 |









#93  
 1,2,4-Trichlorobenzene  
 Concen: 22.317 ug/l  
 RT: 15.02 min Scan# 2178  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20




Instrument : MSVOA\_Y  
 ClientSampled :


| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 180     | 62477 |       |       |
| 182     | 94.5  | 46.9  | 140.8 |
| 145     | 30.7  | 16.0  | 48.0  |



#94  
 Hexachlorobutadiene  
 Concen: 21.495 ug/l  
 RT: 15.12 min Scan# 2195  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 225     | 30315 |       |       |
| 223     | 62.4  | 31.1  | 93.3  |
| 227     | 63.4  | 31.3  | 93.9  |









#95  
 Naphthalene  
 Concen: 21.241 ug/l  
 RT: 15.25 min Scan# 2216  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

Instrument : MSVOA\_Y  
 ClientSampled :

| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 128     | 138436 |       |       |
| 127     | 12.8   | 10.2  | 15.4  |
| 129     | 11.1   | 8.5   | 12.7  |



#96  
 1,2,3-Trichlorobenzene  
 Concen: 22.171 ug/l  
 RT: 15.43 min Scan# 2246  
 Delta R.T. -0.01 min  
 Lab File: VY001293.D  
 Acq: 20 Jan 2020 14:20

| Tgt Ion | Resp  | Lower | Upper |
|---------|-------|-------|-------|
| 180     | 54700 |       |       |
| 182     | 94.0  | 47.5  | 142.5 |
| 145     | 32.8  | 17.1  | 51.3  |

