


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\VY012020\
 Data File : VY001289.D
 Acq On : 20 Jan 2020 11:55
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y011020S.M
 Title : SW846 8260
 Last Update : Tue Jan 21 06:54:04 2020



AutoFind: Scans 1764, 1765, 1766; Background Corrected with Scan 1756

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	14252	PASS
75	95	30	60	50.0	39269	PASS
95	95	100	100	100.0	78512	PASS
96	95	5	9	6.8	5335	PASS
173	174	0.00	2	0.7	445	PASS
174	95	50	100	76.9	60352	PASS
175	174	5	9	6.5	3916	PASS
176	174	95	101	97.0	58528	PASS
177	176	5	9	7.2	4185	PASS