

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012023\
 Data File : VY012295.D
 Acq On : 20 Jan 2023 17:00
 Operator : KP/MD
 Sample : 01232-05
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-P-18B

Quant Time: Jan 23 02:22:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	205706	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	345925	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	328400	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	154191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	115508	69.258	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	138.520%
35) Dibromofluoromethane	7.716	113	101521	55.425	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.860%
50) Toluene-d8	10.179	98	403901	64.936	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	129.880%
62) 4-Bromofluorobenzene	12.483	95	135818	52.600	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.200%
Target Compounds						
40) Benzene	8.161	78	15186	1.690	ug/l	99
52) Toluene	10.246	92	6148	1.028	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012023\
Data File : VY012295.D
Acq On : 20 Jan 2023 17:00
Operator : KP/MD
Sample : 01232-05
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-P-18B

Quant Time: Jan 23 02:22:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

