

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012023\
 Data File : VY012305.D
 Acq On : 20 Jan 2023 20:53
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 23 02:26:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.437	0.362	17.2	95	0.00
3 P	Chloromethane	0.403	0.331	17.9	88	0.00
4 C	Vinyl Chloride	0.471	0.415	11.9#	94	0.00
5 T	Bromomethane	0.364	0.335	8.0	98	0.00
6 T	Chloroethane	0.307	0.288	6.2	98	0.00
7 T	Trichlorofluoromethane	0.919	0.873	5.0	102	0.00
8 T	Diethyl Ether	0.273	0.260	4.8	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.447	0.436	2.5	105	0.00
10 T	Methyl Iodide	0.680	0.622	8.5	89	0.00
11 T	Tert butyl alcohol	0.057	0.047	17.5	100	0.00
12 CM	1,1-Dichloroethene	0.459	0.447	2.6#	100	0.00
13 T	Acrolein	0.023	0.012	47.8#	71	0.00
14 T	Allyl chloride	0.568	0.532	6.3	96	0.00
15 T	Acrylonitrile	0.119	0.119	0.0	99	0.00
16 T	Acetone	0.139	0.108	22.3	69	0.00
17 T	Carbon Disulfide	1.420	1.330	6.3	97	0.00
18 T	Methyl Acetate	0.312	0.307	1.6	100	0.00
19 T	Methyl tert-butyl Ether	1.280	1.271	0.7	97	0.00
20 T	Methylene Chloride	0.762	0.566	25.7#	97	0.00
21 T	trans-1,2-Dichloroethene	0.512	0.507	1.0	100	0.00
22 T	Diisopropyl ether	1.132	1.114	1.6	100	0.00
23 T	Vinyl Acetate	0.705	0.681	3.4	95	0.00
24 P	1,1-Dichloroethane	0.785	0.780	0.6	101	0.00
25 T	2-Butanone	0.154	0.137	11.0	85	0.00
26 T	2,2-Dichloropropane	0.852	0.799	6.2	97	0.00
27 T	cis-1,2-Dichloroethene	0.563	0.573	-1.8	103	0.00
28 T	Bromochloromethane	0.171	0.155	9.4	96	0.00
29 T	Tetrahydrofuran	0.084	0.083	1.2	96	0.00
30 C	Chloroform	0.927	0.941	-1.5#	103	0.00
31 T	Cyclohexane	0.729	0.683	6.3	106	0.00
32 T	1,1,1-Trichloroethane	0.918	0.909	1.0	101	0.00
33 S	1,2-Dichloroethane-d4	0.465	0.405	12.9	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.265	0.244	7.9	99	0.00
36 T	1,1-Dichloropropene	0.450	0.442	1.8	102	0.00
37 T	Ethyl Acetate	0.198	0.189	4.5	94	0.00
38 T	Carbon Tetrachloride	0.581	0.579	0.3	103	0.00
39 T	Methylcyclohexane	0.550	0.560	-1.8	107	0.00
40 TM	Benzene	1.299	1.296	0.2	103	0.00
41 T	Methacrylonitrile	0.117	0.123	-5.1	100	0.00
42 TM	1,2-Dichloroethane	0.413	0.415	-0.5	100	0.00
43 T	Isopropyl Acetate	0.369	0.357	3.3	95	0.00
44 TM	Trichloroethene	0.388	0.392	-1.0	103	0.00
45 C	1,2-Dichloropropane	0.271	0.275	-1.5#	104	0.00
46 T	Dibromomethane	0.181	0.186	-2.8	102	0.00
47 T	Bromodichloromethane	0.457	0.468	-2.4	103	0.00
48 T	Methyl methacrylate	0.170	0.169	0.6	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012023\
 Data File : VY012305.D
 Acq On : 20 Jan 2023 20:53
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 23 02:26:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	97	0.00
50 S	Toluene-d8	1.044	0.895	14.3	99	0.00
51 T	4-Methyl-2-Pentanone	0.197	0.194	1.5	96	0.00
52 CM	Toluene	0.865	0.880	-1.7#	103	0.00
53 T	t-1,3-Dichloropropene	0.478	0.481	-0.6	99	0.00
54 T	cis-1,3-Dichloropropene	0.514	0.512	0.4	99	0.00
55 T	1,1,2-Trichloroethane	0.258	0.261	-1.2	100	0.00
56 T	Ethyl methacrylate	0.338	0.353	-4.4	100	0.00
57 T	1,3-Dichloropropane	0.421	0.431	-2.4	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.112	0.104	7.1	94	0.00
59 T	2-Hexanone	0.150	0.141	6.0	88	0.00
60 T	Dibromochloromethane	0.342	0.360	-5.3	103	0.00
61 T	1,2-Dibromoethane	0.249	0.256	-2.8	102	0.00
62 S	4-Bromofluorobenzene	0.373	0.339	9.1	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.456	0.492	-7.9	110	0.00
65 PM	Chlorobenzene	1.023	1.021	0.2	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.386	0.392	-1.6	103	0.00
67 C	Ethyl Benzene	1.828	1.818	0.5#	103	0.00
68 T	m/p-Xylenes	0.723	0.726	-0.4	103	0.00
69 T	o-Xylene	0.690	0.694	-0.6	103	0.00
70 T	Styrene	1.154	1.173	-1.6	103	0.00
71 P	Bromoform	0.247	0.255	-3.2	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.605	3.627	-0.6	105	0.00
74 T	N-amyl acetate	0.707	0.651	7.9	92	0.00
75 P	1,1,2,2-Tetrachloroethane	0.568	0.546	3.9	98	0.00
76 T	1,2,3-Trichloropropane	0.460	0.505	-9.8	101	0.00
77 T	Bromobenzene	0.868	0.876	-0.9	104	0.00
78 T	n-propylbenzene	4.219	4.114	2.5	103	0.00
79 T	2-Chlorotoluene	2.385	2.352	1.4	103	0.00
80 T	1,3,5-Trimethylbenzene	3.134	3.122	0.4	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.223	0.213	4.5	95	0.00
82 T	4-Chlorotoluene	2.506	2.476	1.2	104	0.00
83 T	tert-Butylbenzene	2.732	2.691	1.5	105	0.00
84 T	1,2,4-Trimethylbenzene	3.091	3.099	-0.3	103	0.00
85 T	sec-Butylbenzene	3.889	3.834	1.4	104	0.00
86 T	p-Isopropyltoluene	3.398	3.372	0.8	105	0.00
87 T	1,3-Dichlorobenzene	1.725	1.698	1.6	103	0.00
88 T	1,4-Dichlorobenzene	1.721	1.673	2.8	103	0.00
89 T	n-Butylbenzene	2.945	2.837	3.7	102	0.00
90 T	Hexachloroethane	0.613	0.610	0.5	106	0.00
91 T	1,2-Dichlorobenzene	1.540	1.514	1.7	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.117	0.113	3.4	98	0.00
93 T	1,2,4-Trichlorobenzene	0.987	0.940	4.8	102	0.00
94 T	Hexachlorobutadiene	0.594	0.582	2.0	108	0.00
95 T	Naphthalene	1.947	1.880	3.4	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012023\
Data File : VY012305.D
Acq On : 20 Jan 2023 20:53
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jan 23 02:26:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.849	0.805	5.2	102	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6