

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012023\
 Data File : VY012289.D
 Acq On : 20 Jan 2023 14:39
 Operator : KP/MD
 Sample : 01190-11
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PE-10

Quant Time: Jan 23 02:20:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	221658	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	368157	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	353670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	166527	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	120338	66.820	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.640%
35) Dibromofluoromethane	7.716	113	110564	56.717	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.440%
50) Toluene-d8	10.179	98	435730	65.888	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	131.780%
62) 4-Bromofluorobenzene	12.483	95	147788	53.779	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.560%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012023\
Data File : VY012289.D
Acq On : 20 Jan 2023 14:39
Operator : KP/MD
Sample : 01190-11
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PE-10

Quant Time: Jan 23 02:20:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY012289.D
Acq: 20 Jan 2023 14:39




Instrument :
MSVOA_Y
ClientSampleId :
PE-10


Tgt Ion:168 Resp: 221658
Ion Ratio Lower Upper
168 100
99 54.0 44.0 66.0



#33
1,2-Dichloroethane-d4
Concen: 66.820 ug/l
RT: 8.142 min Scan# 1037
Delta R.T. -0.001 min
Lab File: VY012289.D
Acq: 20 Jan 2023 14:39

Tgt Ion: 65 Resp: 120338
Ion Ratio Lower Upper
65 100
67 51.4 0.0 103.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY012289.D

Acq: 20 Jan 2023 14:39

Instrument :

MSVOA_Y

ClientSampleId :

PE-10





Tgt Ion:114 Resp: 368157

Ion Ratio Lower Upper

114 100

63 17.8 0.0 39.6

88 15.7 0.0 32.2



#35

Dibromofluoromethane

Concen: 56.717 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY012289.D

Acq: 20 Jan 2023 14:39




Tgt Ion:113 Resp: 110564


Ion Ratio Lower Upper

113 100

111 103.4 83.0 124.4

192 22.0 16.0 24.0








#50
Toluene-d8
Concen: 65.888 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY012289.D
Acq: 20 Jan 2023 14:39




Instrument :
MSVOA_Y
ClientSampleId :
PE-10


Tgt Ion: 98 Resp: 435730
Ion Ratio Lower Upper
98 100
100 64.8 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 53.779 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY012289.D
Acq: 20 Jan 2023 14:39

Tgt Ion: 95 Resp: 147788
Ion Ratio Lower Upper
95 100
174 92.2 0.0 172.0
176 89.3 0.0 167.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.489 min Scan# 11

Delta R.T. -0.001 min

Lab File: VY012289.D

Acq: 20 Jan 2023 14:39

Instrument :

MSVOA_Y

ClientSampleId :

PE-10





Tgt Ion:117 Resp: 353670

Ion Ratio Lower Upper

117 100

82 52.6 44.8 67.2

119 32.0 25.4 38.2



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.428 min Scan# 1904

Delta R.T. 0.006 min

Lab File: VY012289.D

Acq: 20 Jan 2023 14:39

Tgt Ion:152 Resp: 166527

Ion Ratio Lower Upper

152 100

115 56.3 29.1 87.3

150 156.5 0.0 342.6

