

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012025\
 Data File : VY020876.D
 Acq On : 20 Jan 2025 13:01
 Operator : SY/MD
 Sample : Q1132-04
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-2-011725

Quant Time: Jan 21 00:40:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 01:31:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	1250	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.628	114	1440	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.432	117	431	50.000	ug/l	# 0.01
72) 1,4-Dichlorobenzene-d4	13.359	152	184	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	600	56.473	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 112.940%	
35) Dibromofluoromethane	7.640	113	532	58.970	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 117.940%	
50) Toluene-d8	10.115	98	762	24.230	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 48.460%#	
62) 4-Bromofluorobenzene	12.420	95	84	7.233	ug/l	0.01
Spiked Amount	50.000	Range	29 - 146	Recovery	= 14.460%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012025\
Data File : VY020876.D
Acq On : 20 Jan 2025 13:01
Operator : SY/MD
Sample : Q1132-04
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-2-011725

Quant Time: Jan 21 00:40:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 01:31:46 2025
Response via : Initial Calibration

