

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012025\
 Data File : VY020877.D
 Acq On : 20 Jan 2025 13:25
 Operator : SY/MD
 Sample : Q1134-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-1-011725

Quant Time: Jan 21 00:41:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 01:31:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	10192	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.628	114	12973	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	6649	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	3391	39.144	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	78.280%
35) Dibromofluoromethane	7.640	113	4028	49.560	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.120%
50) Toluene-d8	10.115	98	12211	43.099	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	86.200%
62) 4-Bromofluorobenzene	12.420	95	1445	13.811	ug/l	0.01
Spiked Amount	50.000	Range	29 - 146	Recovery	=	27.620%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012025\
Data File : VY020877.D
Acq On : 20 Jan 2025 13:25
Operator : SY/MD
Sample : Q1134-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-1-011725

Quant Time: Jan 21 00:41:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 01:31:46 2025
Response via : Initial Calibration

