

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012025\
 Data File : VY020886.D
 Acq On : 20 Jan 2025 17:20
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jan 21 00:44:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 01:31:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	176524	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	307215	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	270757	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	105633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	95895	63.913	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 127.820%	
35) Dibromofluoromethane	7.640	113	101348	52.657	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.320%	
50) Toluene-d8	10.115	98	362909	54.089	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 108.180%	
62) 4-Bromofluorobenzene	12.414	95	110657	44.661	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 89.320%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012025\
Data File : VY020886.D
Acq On : 20 Jan 2025 17:20
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jan 21 00:44:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 01:31:46 2025
Response via : Initial Calibration

