

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012025\
 Data File : VY020887.D
 Acq On : 20 Jan 2025 17:43
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jan 21 00:44:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 01:31:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	181865	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	309544	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	276161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	110029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	95185	61.577	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.160%	
35) Dibromofluoromethane	7.640	113	101172	52.170	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.340%	
50) Toluene-d8	10.115	98	364988	53.990	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 107.980%	
62) 4-Bromofluorobenzene	12.414	95	112319	44.991	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 89.980%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012025\
Data File : VY020887.D
Acq On : 20 Jan 2025 17:43
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jan 21 00:44:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 01:31:46 2025
Response via : Initial Calibration

