

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012224\
 Data File : VY017073.D
 Acq On : 22 Jan 2024 14:19
 Operator : SY/MD
 Sample : P1207-01
 Misc : 7.64g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EC68-SOUTH

Quant Time: Jan 23 01:45:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 02 14:25:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	99604	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.709	114	207340	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	201084	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	72758	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	44215	44.914	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.820%
35) Dibromofluoromethane	7.740	113	49573	37.520	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	75.040%
50) Toluene-d8	10.197	98	185257	40.718	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	81.440%
62) 4-Bromofluorobenzene	12.495	95	56999	37.132	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	74.260%
Target Compounds						
20) Methylene Chloride	4.747	84	19485	9.273	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012224\
Data File : VY017073.D
Acq On : 22 Jan 2024 14:19
Operator : SY/MD
Sample : P1207-01
Misc : 7.64g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EC68-SOUTH

Quant Time: Jan 23 01:45:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010224S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 02 14:25:32 2024
Response via : Initial Calibration

