



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY012224\  
Data File : VY017064.D  
Acq On : 22 Jan 2024 08:44  
Operator : SY/MD  
Sample : BFB  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y010224S.M  
Title : SW846 8260  
Last Update : Tue Jan 02 14:25:32 2024



AutoFind: Scans 1753, 1754, 1755; Background Corrected with Scan 1745

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 19.4         | 13419      | PASS                |
| 75             | 95              | 30              | 60              | 51.5         | 35701      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 69280      | PASS                |
| 96             | 95              | 5               | 9               | 6.4          | 4411       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.2          | 654        | PASS                |
| 174            | 95              | 50              | 100             | 81.4         | 56416      | PASS                |
| 175            | 174             | 5               | 9               | 7.8          | 4381       | PASS                |
| 176            | 174             | 95              | 101             | 95.5         | 53872      | PASS                |
| 177            | 176             | 5               | 9               | 6.8          | 3670       | PASS                |