


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv012320\
 Data File : VY001343.D
 Acq On : 23 Jan 2020 09:49
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y012120S.M
 Title : SW846 8260
 Last Update : Thu Jan 23 01:21:47 2020



AutoFind: Scans 1764, 1765, 1766; Background Corrected with Scan 1755

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	16013	PASS
75	95	30	60	49.7	41842	PASS
95	95	100	100	100.0	84258	PASS
96	95	5	9	7.2	6062	PASS
173	174	0.00	2	0.9	599	PASS
174	95	50	100	76.2	64186	PASS
175	174	5	9	7.2	4608	PASS
176	174	95	101	95.8	61493	PASS
177	176	5	9	7.1	4344	PASS