

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012323\
 Data File : VY012311.D
 Acq On : 23 Jan 2023 14:14
 Operator : KP/MD
 Sample : 01232-04
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-P-18A

Quant Time: Jan 23 15:16:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	170252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	283949	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	275020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	125590	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	103789	75.557	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 151.120%	
35) Dibromofluoromethane	7.722	113	83042	55.232	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.460%	
50) Toluene-d8	10.179	98	334138	65.483	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 130.960%	
62) 4-Bromofluorobenzene	12.483	95	113735	53.662	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 107.320%	
Target Compounds						
						Qvalue
16) Acetone	3.948	43	2855	6.022	ug/l	96
25) 2-Butanone	6.966	43	3110	5.942	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012323\
Data File : VY012311.D
Acq On : 23 Jan 2023 14:14
Operator : KP/MD
Sample : 01232-04
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-P-18A

Quant Time: Jan 23 15:16:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

