

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012323\
 Data File : VY012321.D
 Acq On : 23 Jan 2023 18:08
 Operator : KP/MD
 Sample : 01251-12
 Misc : 3.64g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOT-VOC-4

Quant Time: Jan 24 01:30:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	67771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	102005	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	84525	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	32509	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	31874	57.316	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 114.640%	
35) Dibromofluoromethane	7.716	113	28104	52.033	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.060%	
50) Toluene-d8	10.179	98	118557	64.618	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 129.240%	
62) 4-Bromofluorobenzene	12.483	95	29199	38.349	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 76.700%	
Target Compounds						
16) Acetone	3.942	43	3355	17.778	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012323\
Data File : VY012321.D
Acq On : 23 Jan 2023 18:08
Operator : KP/MD
Sample : 01251-12
Misc : 3.64g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOT-VOC-4

Quant Time: Jan 24 01:30:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

