

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012323\
 Data File : VY012324.D
 Acq On : 23 Jan 2023 19:18
 Operator : KP/MD
 Sample : 01253-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-04-012023

Quant Time: Jan 24 01:30:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	126165	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	212965	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	209744	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	96063	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	83330	82.217	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	164.440%#
35) Dibromofluoromethane	7.722	113	64148	56.886	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.780%
50) Toluene-d8	10.179	98	256392	67.104	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	134.200%
62) 4-Bromofluorobenzene	12.483	95	85640	53.874	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012323\
Data File : VY012324.D
Acq On : 23 Jan 2023 19:18
Operator : KP/MD
Sample : 01253-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-04-012023

Quant Time: Jan 24 01:30:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

