

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012323\
 Data File : VY012327.D
 Acq On : 23 Jan 2023 20:29
 Operator : KP/MD
 Sample : 01243-05
 Misc : 6.60g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TJRC-011923-B4-20-30

Quant Time: Jan 24 01:31:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	157813	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	267876	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	264002	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	127447	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	101297	79.781	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 159.560%	
35) Dibromofluoromethane	7.716	113	79807	56.265	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 112.540%	
50) Toluene-d8	10.179	98	319406	66.415	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 132.820%	
62) 4-Bromofluorobenzene	12.483	95	110082	55.054	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 110.100%	
Target Compounds						
					Qvalue	
30) Chloroform	7.502	83	3065	1.048	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012323\
Data File : VY012327.D
Acq On : 23 Jan 2023 20:29
Operator : KP/MD
Sample : 01243-05
Misc : 6.60g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TJRC-011923-B4-20-30

Quant Time: Jan 24 01:31:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

