

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012422\
 Data File : VY007277.D
 Acq On : 24 Jan 2022 18:32
 Operator : SY/MD
 Sample : N1245-04
 Misc : 4.84g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-2(2.5-3)

Quant Time: Jan 25 00:34:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 14 05:48:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	86389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	161286	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	155747	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	68991	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	56716	53.505	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.020%
35) Dibromofluoromethane	7.722	113	54090	49.597	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.200%
50) Toluene-d8	10.185	98	201238	50.499	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.000%
62) 4-Bromofluorobenzene	12.483	95	72595	51.213	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012422\
Data File : VY007277.D
Acq On : 24 Jan 2022 18:32
Operator : SY/MD
Sample : N1245-04
Misc : 4.84g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-2(2.5-3)

Quant Time: Jan 25 00:34:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011322S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 14 05:48:29 2022
Response via : Initial Calibration

