

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012423\
 Data File : VY012336.D
 Acq On : 24 Jan 2023 12:55
 Operator : KP/MD
 Sample : 01243-07
 Misc : 6.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TJRC-011923-B4-30-40

Quant Time: Jan 25 01:31:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	163124	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	270242	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	260428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	120931	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	100261	76.213	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	152.420%
35) Dibromofluoromethane	7.722	113	80805	56.470	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.940%
50) Toluene-d8	10.179	98	324347	66.883	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	133.760%
62) 4-Bromofluorobenzene	12.483	95	106751	52.921	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012423\
Data File : VY012336.D
Acq On : 24 Jan 2023 12:55
Operator : KP/MD
Sample : 01243-07
Misc : 6.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TJRC-011923-B4-30-40

Quant Time: Jan 25 01:31:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

