

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012423\
 Data File : VY012342.D
 Acq On : 24 Jan 2023 15:15
 Operator : KP/MD
 Sample : 01260-10
 Misc : 4.71g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-24

Quant Time: Jan 25 01:32:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	155925	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	261044	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	257720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	122265	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	94253	74.883	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	149.760%
35) Dibromofluoromethane	7.722	113	77405	56.000	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.000%
50) Toluene-d8	10.179	98	314147	67.075	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	134.140%
62) 4-Bromofluorobenzene	12.483	95	103899	53.322	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012423\
Data File : VY012342.D
Acq On : 24 Jan 2023 15:15
Operator : KP/MD
Sample : 01260-10
Misc : 4.71g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-24

Quant Time: Jan 25 01:32:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

