

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012423\
 Data File : VY012347.D
 Acq On : 24 Jan 2023 17:12
 Operator : KP/MD
 Sample : 01253-01
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-04-012023

Quant Time: Jan 25 01:32:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 04:31:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	148369	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	254731	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	247163	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	117522	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	91830	76.776	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	153.560%
35) Dibromofluoromethane	7.716	113	74383	55.147	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.300%
50) Toluene-d8	10.179	98	306775	67.127	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	134.260%
62) 4-Bromofluorobenzene	12.483	95	102481	53.898	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY012423\
Data File : VY012347.D
Acq On : 24 Jan 2023 17:12
Operator : KP/MD
Sample : 01253-01
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-04-012023

Quant Time: Jan 25 01:32:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y011623S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 04:31:47 2023
Response via : Initial Calibration

